

COPY

EXHIBIT G-2  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY  
STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000  
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST S. Bickley AGENCY Yavapai County Sheriff's Office  
DATE 11-09-15 TIME 12:05:57  
INTOXILYZER SERIAL # 80-006228 LOCATION DUI Truck

1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0.100 AC.  
OR  
Pour a standard alcohol concentration solution 0.100 AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach  $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
4. Air blank completed.
5. Calibration check completed. Test results 0.096 AC.
6. Air blank completed.
7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE S. Bickley 66861/1142

INTOXILYZER 8000  
Location: YAVAPAI COUNTY SO  
Serial Number: 80-006228  
Core Version: 8105.48  
11/09/2015 12:05:57

-----  
Standard Lot#: 01314100A3  
Last Changed By: J. MURRAY5597 #N56  
-----

QAS: S. BLAKLEY SR #L6861  
YAVAPAI CSO  
-----

Operator: S. BLAKLEY SR #L6861  
YAVAPAI CSO  
-----

Test	g/210L	Time
Air Blank	0.000	12:06:42
0.100 Cal Check	0.096	12:07:02
Air Blank	0.000	12:07:32

Preventative Maintenance  
Performed -- Timer Reset

EXHIBIT G-4  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST S. Blakley AGENCY Maricopa County Sheriff's Office  
DATE 11-09-15 TIME \_\_\_\_\_  
INTOXILYZER SERIAL # 80-006228 LOCATION DUI Truck

(X) 1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

- (X) 1. Clock time check.  
(X) 2. Date check.

OPERATIONAL TESTS

- (X) 1. Alcohol-free subject test result 0.000 AC.  
(X) 2. Error recognition logic system functioning.  
Not a Successfully Completed Test Sequence printed  
(X) 3. Proper sample recognition system.  
Not a Successfully Completed Test Sequence printed  
Deficient sample printed.  
(X) 4. Standard Calibration Check standard 0.100 AC. Result 0.096 AC.

Instrument is operating properly and accurately. YES  NO \_\_\_\_\_

COMMENTS

Bottle S/N C1314100A3 Exp 02-11-2016  
\_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_

SIGNATURE S. Blakley L6861/N42

INTOXILYZER 8000  
Location: YAVAPAI COUNTY SO  
Serial Number: 80-006228  
Core Version: 8105.48  
11/09/2015

12:09:25

---

Standard Lot#: 01314100A3  
Last Changed By: J. MURRAY5597 #N56

---

QAS: S. BLAKLEY SR #L6861  
YAVAPAI CSO

---

Operator: S. BLAKLEY SR #L6861  
YAVAPAI CSO

---

Subject: SQAP,TEST  
DOB: 02/25/1969  
Sex: M Weight: 210

---

15 Minute Deprivation Period? Yes

---

Test	g/210L	Time
Air Blank	0.000	12:11:07
Diagnostic Test	Pass	12:11:42
Air Blank	0.000	12:12:09
0.100 Cal Check	0.096	12:12:30
Air Blank	0.000	12:13:00
Subject Test	0.000	12:13:22
Air Blank	ABT*	12:13:40
Air Blank	IPS**	12:13:50

\*Sequence Aborted

\*\*Improper Sample

Not a Successfully Completed Test Sequence

INTOXILYZER 8000  
Location: YAVAPAI COUNTY SO  
Serial Number: 80-006228  
Core Version: 8105.48  
11/09/2015

12:14:33

---

Standard Lot#: 01314100A3  
Last Changed By: J. MURRAY5597 #N56

---

QAS: S. BLAKLEY SR #L6861  
YAVAPAI CSO

---

Operator: S. BLAKLEY SR #L6861  
YAVAPAI CSO

---

Subject: SQAAP, TEST  
DOB: 02/25/1969  
Sex: M Weight: 210

---

15 Minute Deprivation Period? Yes

---

Test	g/210L	Time
Air Blank	0.000	12:15:48
Diagnostic Test	Pass	12:16:23
Air Blank	0.000	12:16:50
0.100 Cal Check	0.097	12:17:11
Air Blank	0.000	12:17:40
Subject Test	0.033*	12:20:48
Air Blank	0.000	12:21:19
Air Blank	0.000	12:21:46
Subject Test	RGE**	12:22:06
Air Blank	0.000	12:22:41

\*Deficient Sample

\*\*Range Exceeded

Not a Successfully Completed Test Sequence

INTOXILYZER 8000  
Location: YAVAPAI COUNTY SO  
Serial Number: 80-006228  
Core Version: 8105.48  
11/09/2015 12:23:18

---

Standard Lot#: 01314100A3  
Last Changed By: J. MURRAY5597 #N56

---

QAS: S. BLAKLEY SR #L6861  
YAVAPAI CSO

---

Operator: S. BLAKLEY SR #L6861  
YAVAPAI CSO

---

Subject: SQAP,TEST  
DOB: 02/25/1969  
Sex: M Weight: 210

---

15 Minute Deprivation Period? Yes

---

Test	g/210L	Time
Air Blank	0.000	12:24:19
Diagnostic Test	RFI*	12:24:51
Air Blank	0.000	12:25:19

\* RFI Detect

Not a Successfully Completed Test Sequence