

EXHIBIT G-2
 THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)
 ARIZONA DEPARTMENT OF PUBLIC SAFETY
 STANDARD QUALITY ASSURANCE PROCEDURES
 INTOXILYZER MODEL 8000
 STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST Scott Blakley Sr. AGENCY Yavapai County SO
 DATE 1-24-18 TIME 20:48:14
 INTOXILYZER SERIAL # 80-006228 LOCATION DUI Truck

- (X) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0.100 AC.
 OR
 Pour a standard alcohol concentration solution NA AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C
- (N) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (X) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (X) 4. Air blank completed.
- (X) 5. Calibration check completed. Test results 0.100 AC.
- (X) 6. Air blank completed.
- (X) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE Scott Blakley Sr.

Tank Serial # 06717100A5
 Exp 5-5-2019
 Pres 1000 Psi

Past 31 day

DR 18-008965

INTOXILYZER 8000

Location: YAVAPAI COUNTY SO

Serial Number: 80-006228

Core Version: 8105.48

01/24/2018

20:48:14

Standard Lot#: 06717100A5

Last Changed By: S. BLAKLEY SR #L6861

QAS: S. BLAKLEY SR #L6861

YAVAPAI CSO

Operator: S. BLAKLEY SR #L6861

YAVAPAI CSO

Test	g/210L	Time
Air Blank	0.000	20:48:45
0.100 Cal Check	0.100	20:49:06
Air Blank	0.000	20:49:35

Preventative Maintenance
Performed -- Timer Reset

INTOXILYZER 8000

Location: YAVAPAI COUNTY SO

Serial Number: 80-006228

Core Version: 8105.48

01/24/2018

20:49:46

Standard Lot#: 06717100A5

Last Changed By: S. BLAKLEY SR #L6861

QAS: S. BLAKLEY SR #L6861

YAVAPAI CSO

Operator: S. BLAKLEY SR #L6861

YAVAPAI CSO

Test	g/210L	Time
Air Blank	0.000	20:50:20
0.100 Cal Check	0.100	20:50:41
Air Blank	0.000	20:51:10

Preventative Maintenance
Performed -- Timer Reset

Temp Reg Fail

DR 18-002965

INTOXILYZER 8000

Location: YAVAPAI COUNTY SO

Serial Number: 80-006228

Core Version: 8105.48

01/24/2018

20:55:17

Standard Lot#: 06717100A5

Last Changed By: S. BLAKLEY SR #L6861

QAS: S. BLAKLEY SR #L6861

YAVAPAI CSO

Operator: S. BLAKLEY SR #L6861

YAVAPAI CSO

Test	g/210L	Time
Air Blank	0.000	20:55:49
0.100 Cal Check	0.100	20:56:10
Air Blank	0.000	20:56:39

Preventative Maintenance
Performed -- Timer Reset

EXHIBIT G-4
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST Scott Blakley Sr. AGENCY Yavapai County SO
DATE 1-24-18 TIME 21:00:21
INTOXILYZER SERIAL # 80-006228 LOCATION YCSO DUI Truck

() 1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

- () 1. Clock time check.
- () 2. Date check.

OPERATIONAL TESTS

- () 1. Alcohol-free subject test result 0. 000 AC.
- () 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
- () 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
- () 4. Standard Calibration Check standard 0. 100 AC. Result 0. 099 AC.

Instrument is operating properly and accurately. YES NO

COMMENTS

SIGNATURE Scott Blakley Sr.

. DR 18-002965

INTOXILYZER 8000

Location: YAVAPAI COUNTY SO

Serial Number: 80-006228

Core Version: 8105.48

01/24/2018

21:00:21

Standard Lot#: 06717100A5

Last Changed By: S. BLAKLEY SR #L6861

QAS: S. BLAKLEY SR #L6861
YAVAPAI CSO

Operator: S. BLAKLEY SR #L6861
YAVAPAI CSO

Subject: SQAP,TEST

DOB: 10/20/1960

Sex: F Weight: 115

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	21:01:50
Diagnostic Test	Pass	21:02:25
Air Blank	0.000	21:02:53
0.100 Cal Check	0.099	21:03:13
Air Blank	0.000	21:03:43
Subject Test	0.000	21:04:02
Air Blank	0.000	21:04:29
Wait	ABT*	
Air Blank	0.000	21:05:08

*Sequence Aborted

Not a Successfully Completed Test Sequence

DR 18-002965

INTOXILYZER 8000
Location: YAVAPAI COUNTY SO
Serial Number: 80-006228
Core Version: 8105.48
01/24/2018 21:06:00

Standard Lot#: 06717100A5
Last Changed By: S. BLAKLEY SR #L6861

QAS: S. BLAKLEY SR #L6861
YAVAPAI CSO

Operator: S. BLAKLEY SR #L6861
YAVAPAI CSO

Subject: SQAP, TEST
DOB: 10/20/1960
Sex: F Weight: 115

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	IPS*	21:06:43
Air Blank	RFI**	21:06:56

*Improper Sample
**RFI Detect

Not a Successfully Completed Test Sequence

DR 18-002965

INTOXILYZER 8000

Location: YAVAPAI COUNTY SO

Serial Number: 80-006228

Core Version: 8105.48

01/24/2018

21:07:29

Standard Lot#: 06717100A5

Last Changed By: S. BLAKLEY SR #L6861

QAS: S. BLAKLEY SR #L6861
YAVAPAI CSO

Operator: S. BLAKLEY SR #L6861
YAVAPAI CSO

Subject: SQAP,TEST

DOB: 10/20/1960

Sex: F

Weight: 115

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	21:08:23
Diagnostic Test	Pass	21:08:58
Air Blank	0.000	21:09:25
0.100 Cal Check	0.099	21:09:46
Air Blank	0.000	21:10:15
Subject Test	0.000*	21:13:23
Air Blank	0.000	21:13:54
Air Blank	0.000	21:14:22
Subject Test	RGE**	21:14:45
Air Blank	0.000	21:15:23

*Deficient Sample

**Range Exceeded

Not a Successfully Completed Test Sequence

DR 18-002965

INTOXILYZER 8000

Location: YAVAPAI COUNTY SO

Serial Number: 80-006228

Core Version: 8105.48

01/24/2018

21:16:08

Standard Lot#: 06717100A5

Last Changed By: S. BLAKLEY SR #L6861

QAS: S. BLAKLEY SR #L6861

YAVAPAI CSO

Operator: S. BLAKLEY SR #L6861

YAVAPAI CSO

Subject: SQAP,TEST

DOB: 10/20/1960

Sex: F

Weight: 115

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	21:17:02
Diagnostic Test	Pass	21:17:37
Air Blank	0.000	21:18:05
0.100 Cal Check	0.097	21:18:25
Air Blank	0.000	21:18:55
Subject Test	0.038*	21:22:02
Air Blank	0.000	21:22:34
Air Blank	0.000	21:23:02
Subject Test	XXX**	21:23:22
Air Blank	0.000	21:23:56

*Deficient Sample

**Invalid Sample

Begin new deprivation period

Not a Successfully Completed Test Sequence

DR 18-002965

EXHIBIT G-2
 THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)
 ARIZONA DEPARTMENT OF PUBLIC SAFETY
 STANDARD QUALITY ASSURANCE PROCEDURES
 INTOXILYZER MODEL 8000
 STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST Scott Blakley Sr. AGENCY Yavapai County SO
 DATE 1-24-18 TIME 21:25:03
 INTOXILYZER SERIAL # 80-006228 LOCATION YCSO OUT Track

- (x) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0.100 AC.
 OR
 Pour a standard alcohol concentration solution NA AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C
- (x) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (x) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (x) 4. Air blank completed.
- (x) 5. Calibration check completed. Test results 0. 099 AC.
- (x) 6. Air blank completed.
- (x) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE Scott Blakley Sr.

D.R 18-002965

INTOXILYZER 8000

Location: YAVAPAI COUNTY SO

Serial Number: 80-006228

Core Version: 8105.48

01/24/2018

21:25:03

Standard Lot#: 06717100A5

Last Changed By: S. BLAKLEY SR #L6861

QAS: S. BLAKLEY SR #L6861

YAVAPAI CSO

Operator: S. BLAKLEY SR #L6861

YAVAPAI CSO

Test	g/210L	Time
Air Blank	0.000	21:25:33
0.100 Cal Check	0.099	21:25:54
Air Blank	0.000	21:26:23

Preventative Maintenance
Performed -- Timer Reset