

DR 18-011964

COPY

EXHIBIT G-2
 THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)
 ARIZONA DEPARTMENT OF PUBLIC SAFETY
 STANDARD QUALITY ASSURANCE PROCEDURES
 INTOXILYZER MODEL 8000
 STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST Scott Blakley Sr AGENCY Yavapai County SO
 DATE 4-3-18 TIME 21:34:55
 INTOXILYZER SERIAL # 80-006228 LOCATION YCSO DUI Truck

- (X) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0.100 AC.
 OR
 Pour a standard alcohol concentration solution NA AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- (X) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (X) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (X) 4. Air blank completed.
- (X) 5. Calibration check completed. Test results 0.098 AC.
- (X) 6. Air blank completed.
- (X) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE Scott Blakley Sr

DPS Form Exh G-2 (Rev 05-01)

Tank Serial # 06717100A5
 Exp 5-5-2019
 Pres 900 psi

INTOXILYZER 8000
Location: YAVAPAI COUNTY SO
Serial Number: 80-006228
Core Version: 8105.48
04/03/2018 21:29:29

Standard Lot#: 06717100A5
Last Changed By: S. BLAKLEY SR #L6861

QAS: S. BLAKLEY SR #L6861
YAVAPAI CSO

Operator: S. BLAKLEY SR #L6861
YAVAPAI CSO

Test	g/210L	Time
Air Blank	0.000	21:30:00
0.100 Cal Check	0.099	21:30:21
Air Blank	0.000	21:30:50

Preventative Maintenance
Performed -- Timer Reset

Past 31 day

INTOXILYZER 8000

Location: YAVAPAI COUNTY SO

Serial Number: 80-006228

Core Version: 8105.48

04/03/2018

21:34:55

Standard Lot#: 06717100A5

Last Changed By: S. BLAKLEY SR #L6861

QAS: S. BLAKLEY SR #L6861

YAVAPAI CSO

Operator: S. BLAKLEY SR #L6861

YAVAPAI CSO

Test	g/210L	Time
Air Blank	0.000	21:35:27
0.100 Cal Check	0.098	21:35:47
Air Blank	0.000	21:36:17

Preventative Maintenance
Performed -- Timer Reset