

18-031442

Copy

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST Scott Blakley Sr. AGENCY Yavapai County SO

DATE 8-25-18 TIME 01:02:59

INTOXILYZER SERIAL # 80-006228 LOCATION YCSO DUI Truck

- (X) -1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0.106 AC.
OR
Pour a standard alcohol concentration solution N/A AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- (X) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- () 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (X) 4. Air blank completed.
- (X) 5. Calibration check completed. Test results 0.099 AC.
- (X) 6. Air blank completed.
- (X) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE Scott Blakley Sr.

DPS Form Exh G-2 (Rev 05-01)

Tank Serial # 06717100A5

Exp 05-05-2019

PPAS 850PSi

DR18-031442

INTOXILYZER 8000

Location: YAVAPAI COUNTY SO

Serial Number: 80-006228

Core Version: 8105.48

08/25/2018

00:59:18

Standard Lot#: 06717100A5

Last Changed By: S. BLAKLEY SR #L6861

QAS: S. BLAKLEY SR #L6861

YAVAPAI CSO

Operator: S. BLAKLEY SR #L6861

YAVAPAI CSO

Test	g/210L	Time
Air Blank	0.000	00:59:48
0.100 Cal Check	0.099	01:00:09
Air Blank	0.000	01:00:39

Preventative Maintenance
Performed -- Timer Reset

Past 31 day

AIR READ

18-031442

INTOXILYZER 8000

Location: YAVAPAI COUNTY SO

Serial Number: 80-006228

Core Version: 8105.48

08/25/2018

01:02:59

Standard Lot#: 06717100A5

Last Changed By: S. BLAKLEY SR #L6861

QAS: S. BLAKLEY SR #L6861

YAVAPAI CSO

Operator: S. BLAKLEY SR #L6861

YAVAPAI CSO

Test	g/210L	Time
Air Blank	0.000	01:03:28
0.100 Cal Check	0.095	01:03:49
Air Blank	0.000	01:04:18

Preventative Maintenance
Performed -- Timer Reset

DR18-03442

EXHIBIT G-4
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST Scott Blakley Sr AGENCY Yavapai County SO

DATE 8-25-18 TIME 01:06:05

INTOXILYZER SERIAL # 80-06228 LOCATION YCSO DUI Truck

(X) 1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

- (X) 1. Clock time check.
- (X) 2. Date check.

OPERATIONAL TESTS

- (X) 1. Alcohol-free subject test result 0.000 AC.
- (X) 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
- (X) 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
- (X) 4. Standard Calibration Check standard 0.100 AC. Result 0.096 AC.

Instrument is operating properly and accurately. YES X NO

COMMENTS

SIGNATURE Scott Blakley Sr

18-031442

INTOXILYZER 8000
Location: YAVAPAI COUNTY SO
Serial Number: 80-006228
Core Version: 8105.48
08/25/2018

01:06:05

Standard Lot#: 06717100A5
Last Changed By: S. BLAKLEY SR #L6861

QAS: S. BLAKLEY SR #L6861
YAVAPAI CSO

Operator: S. BLAKLEY SR #L6861
YAVAPAI CSO

Subject: SQAP, TEST
DOB: 10/20/1960
Sex: F Weight: 115

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	01:07:08
Diagnostic Test	Pass	01:07:43
Air Blank	0.000	01:08:11
0.100 Cal Check	0.096	01:08:31
Air Blank	0.000	01:09:01
Subject Test	0.000	01:09:16
Air Blank	0.000	01:09:44
Wait	ABT*	
Air Blank	0.000	01:10:20

*Sequence Aborted

Subject: SQAP, TEST

DOB: 10/20/1960

Not a Successfully Completed Test Sequence

18031442

INTOXILYZER 8000
Location: YAVAPAI COUNTY SO
Serial Number: 80-006228
Core Version: 8105.48
08/25/2018 01:11:02

Standard Lot#: 06717100A5
Last Changed By: S. BLAKLEY SR #L6861

QAS: S. BLAKLEY SR #L6861
YAVAPAI CSO

Operator: S. BLAKLEY SR #L6861
YAVAPAI CSO

Subject: SQAP, TEST
DOB: 10/20/1960
Sex: F Weight: 115

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	IPS*	01:11:40
Air Blank	RFI**	01:11:53

8/25/2018
*Improper Sample
**RFI Detect

~~YAVAPAI CSO~~

~~Operator: S. BLAKLEY SR~~
~~YAVAPAI CSO~~

~~DOB: 10/20/1960~~

~~15 Minute Deprivation Period? Yes~~

~~Test~~
~~RFI Detect~~
~~RFI Detect~~

Not a Successfully Completed Test Sequence

18-031442

INTOXILYZER 8000
Location: YAVAPAI COUNTY SO
Serial Number: 80-006228
Core Version: 8105.48
08/25/2018 01:12:33

Standard Lot#: 06717100A5
Last Changed By: S. BLAKLEY SR #L6861

QAS: S. BLAKLEY SR #L6861
YAVAPAI CSO

Operator: S. BLAKLEY SR #L6861
YAVAPAI CSO

Subject: SQAP, TEST
DOB: 10/20/1960
Sex: F Weight: 115

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	01:13:29
Diagnostic Test	Pass	01:14:04
Air Blank	0.000	01:14:31
0.100 Cal Check	0.096	01:14:52
Air Blank	0.000	01:15:22
Subject Test	0.000*	01:18:29
Air Blank	0.000	01:19:01
Air Blank	0.000	01:19:28
Subject Test	XXX**	01:19:49
Air Blank	0.000	01:20:23

*Deficient Sample

**Invalid Sample

Begin new deprivation period

Not a Successfully Completed Test Sequence

Subject Test
Air Blank
Air Blank
Subject Test
Air Blank

18-031442

INTOXILYZER 8000
Location: YAVAPAI COUNTY SO
Serial Number: 80-006228
Core Version: 8105.48
08/25/2018 01:21:03

Standard Lot#: 06717100A5
Last Changed By: S. BLAKLEY SR #L6861

QAS: S. BLAKLEY SR #L6861
YAVAPAI CSO

Operator: S. BLAKLEY SR #L6861
YAVAPAI CSO

Test	g/210L	Time
Air Blank	0.000	01:21:33
0.100 Cal Check	0.096	01:21:54
Air Blank	0.000	01:22:23

Preventative Maintenance
Performed -- Timer Reset

Operator: S. BLAKLEY SR #L6861
YAVAPAI CSO

0.100 Cal Check

Air Blank