



Arizona Department of Public Safety
P. O. Box 6638
Phoenix, AZ 85005-6638

"Courteous Vigilance"

**Attn QAS: Please perform a 90 day SQAP before
placing Intoxilyzer back in service.**

Intoxilyzer SN: 80-00 6230



ARIZONA DEPARTMENT OF PUBLIC SAFETY
INTOXILYZER MAINTENANCE REPORT

AGENCY Prescott Police Department
Prescott, AZ 86303

OFFICER ATTENBERGER, #252

DATE June 19, 2014

CR NO. I800062300001

Page 1 of 1

INSTRUMENT:

80006230

INSTRUMENT LOCATION:

Prescott Police Department

MAINTENANCE REQUESTED:

Preventative Maintenance

MAINTENANCE PERFORMED:

Preventative Maintenance Performed:

Calibration, Verification

A handwritten signature in black ink, appearing to read "Joseph M Slowinski".

JOSEPH M SLOWINSKI, #6101
Electronics Technician
Central Regional Crime Laboratory
P. O. Box 6638
Phoenix, AZ 85005
(602) 223-2394





7 Eastgate Dr. • P.O. Box 790 • Jacksonville, IL 62651-0790
217-245-2183 • Fax: 217-243-7634 • www.ilmoproducts.com

Certificate of Analysis

Certificate ID: 6434
Part #: BAC105L100T
Cylinder Size: 105L
Lot Number: 07014100A1
Expiration: 5/1/2016

0.100 BAC (For the calibration of instruments used to determine breath alcohol concentration)

Contents: 105 Liters @ 1000 psig 70°F (21°C)

Component:	Concentration:	Accuracy:	Method:
Ethanol	260 ppm	+/- 0.002 or 2%	NDIR
Nitrogen	Balance	BAC whichever is greater	

*NIST Standard Reference Material
Cylinder No. CCI4290 / Job No. 09160202
Certified 212.8 µmol/mol Ethanol in Nitrogen
for ILMO Products Co., Jacksonville, IL

Store in dry area, away from sources of heat, ignition
and direct sunlight. Do not allow storage area to
exceed 52 °C (125 °F).


Specialty Gas Lab Tech

04/14/14
Date

Distributed by: CMI Inc.
316 East Ninth Street
Owensboro, KY 42303
Phone 866-835-0690
www.alcoholtest.com



*NEW INTOX. 8000 PUT INTO SERVICE.

252

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST *T. A. ... 252* AGENCY *Prescott PIR*
DATE *06/26/14* TIME *1615*
INTOXILYZER SERIAL # *80-006230* LOCATION *P.P.D.*

- () 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution *-100* AC.
OR
Pour a standard alcohol concentration solution AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- () 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- () 3. Go to the "Control Testing Menu". Select "*D*" for dry control test or "W" for wet control test. After selection is made press ENTER.
- () 4. Air blank completed.
- () 5. Calibration check completed. Test results *0.098* AC.
- () 6. Air blank completed.
- () 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE *[Signature] 252*

INTOXILYZER 8000
Location: PRESCOTT PD
Serial Number: 80-006230
Core Version: 8105.48
06/26/2014 16:15:10

Standard Lot#: 1234567890

Last Changed By: A. A #A

QAS: T. ATTENBERGER #252
PRESCOTT PD

Operator: T. ATTENBERGER #252
PRESCOTT PD

Test	g/210L	Time
Air Blank	0.000	16:15:42
0.100 Cal Check	0.098	16:16:03
Air Blank	0.000	16:16:33

Preventative Maintenance
Performed -- Timer Reset



EXHIBIT G-4
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST T. Anderson 252 AGENCY Prescott P.D.

DATE 062614 TIME 1617

INTOXILYZER SERIAL # 80-006230 LOCATION P.P.D.

() 1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

- () 1. Clock time check.
- () 2. Date check.

OPERATIONAL TESTS

- () 1. Alcohol-free subject test result 0. 060 AC.
- () 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
- () 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
- () 4. Standard Calibration Check standard 0. 100 AC. Result 0. 098 AC.

Instrument is operating properly and accurately. YES NO

COMMENTS

ALL FUNCTIONING CORRECTLY.

SIGNATURE T. Anderson 252

INTOXILYZER 8000
Location: PRESCOTT PD
Serial Number: 80-006230
Core Version: 8105.48
06/26/2014

16:17:35

Standard Lot#: 1234567890
Last Changed By: A. A

#A

QAS: T. ATTENBERGER
PRESCOTT PD

#252

Operator: T. ATTENBERGER
PRESCOTT PD

#252

Subject: SQAP, TEST
DOB: 01/01/1900
Sex: M

Weight: 300

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	16:18:28
Diagnostic Test	Pass	16:19:03
Air Blank	0.000	16:19:31
0.100 Cal Check	0.098	16:19:52
Air Blank	0.000	16:20:21
Subject Test	0.000	16:20:44
Air Blank	0.000	16:21:11
Wait	ABT*	
Air Blank	0.000	16:21:47

*Sequence Aborted



Not a Successfully Completed Test Sequence

INTOXILYZER 8000

Location: PRESCOTT PD

Serial Number: 80-006230

Core Version: 8105.48

06/26/2014

16:22:07

Standard Lot#: 1234567890

Last Changed By: A. A

#A

QAS: T. ATTENBERGER

#252

PRESCOTT PD

Operator: T. ATTENBERGER

#252

PRESCOTT PD

Subject: SQAP, TEST

DOB: 01/01/1900

Sex: M

Weight: 300

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	IPS*	16:22:46
Air Blank	0.000	16:23:14

*Improper Sample



Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: PRESCOTT PD
Serial Number: 80-006230
Core Version: 8105.48
06/26/2014 16:23:32

Standard Lot#: 1234567890
Last Changed By: A. A #A

QAS: T. ATTENBERGER #252
PRESCOTT PD

Operator: T. ATTENBERGER #252
PRESCOTT PD

Subject: SQAP, TEST
DOB: 01/01/1900
Sex: M Weight: 300

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	RFI*	16:24:22
Air Blank	0.000	16:24:50

*RFI Detect

A handwritten signature in black ink, appearing to be 'J. J. 252', is written over the page.

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: PRESCOTT PD
Serial Number: 80-006230
Core Version: 8105.48
06/26/2014

16:25:08

Standard Lot#: 1234567890
Last Changed By: A. A

#A

QAS: T. ATTENBERGER
PRESCOTT PD

#252

Operator: T. ATTENBERGER
PRESCOTT PD

#252

Subject: SQAP, TEST
DOB: 01/01/1900
Sex: M

Weight: 300

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	16:25:58
Diagnostic Test	Pass	16:26:33
Air Blank	0.000	16:27:01
0.100 Cal Check	0.099	16:27:21
Air Blank	0.000	16:27:51
Subject Test	0.021*	16:30:58
Air Blank	0.000	16:31:29
Air Blank	0.000	16:31:56
Subject Test	XXX**	16:32:25
Air Blank	0.000	16:32:57

*Deficient Sample

**Invalid Sample

Begin new deprivation period



Not a Successfully Completed Test Sequence

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST T. Anderson 252 AGENCY Prescott PD

DATE 062614 TIME 1633

INTOXILYZER SERIAL # 80-006230 LOCATION P.P.D

- () 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- () 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- () 3. Go to the "Control Testing Menu". Select D for dry control test or "W" for wet control test. After selection is made press ENTER.
- () 4. Air blank completed.
- () 5. Calibration check completed. Test results 0.098 AC.
- () 6. Air blank completed.
- () 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE T. Anderson 252

INTOXILYZER 8000
Location: PRESCOTT PD
Serial Number: 80-006230
Core Version: 8105.48
06/26/2014 16:33:37

Standard Lot#: 1234567890
Last Changed By: A. A #A

QAS: T. ATTENBERGER #252
PRESCOTT PD

Operator: T. ATTENBERGER #252
PRESCOTT PD

Test	g/210L	Time
Air Blank	0.000	16:34:07
0.100 Cal Check	0.098	16:34:28
Air Blank	0.000	16:34:58

Preventative Maintenance
Performed -- Timer Reset

