

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST S. Heston AGENCY Prescott P.D.
DATE 3/8/16 TIME 0359
INTOXILYZER SERIAL # 80-006230 LOCATION 222 S. Main St

- ✓ 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C
- ✓ 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
✓ 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
✓ 4. Air blank completed.
✓ 5. Calibration check completed. Test results 0. 100 AC.
✓ 6. Air blank completed.
✓ 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE _____
S. Heston

DPS Form Exh G-2 (Rev 05-01)

INTOXILYZER 8000
Location: PRESCOTT PD
Serial Number: 80-006230
Core Version: 8105.48
03/08/2016 03:54:51

Standard Lot#: 25514100A2
Last Changed By: S. HILTON #372

QAS: S. HILTON #372
PRESCOTT PD

Operator: S. HILTON #372
PRESCOTT PD

Test	g/210L	Time
Air Blank	0.000	03:55:24
0.100 Cal Check	0.100	03:55:44
Air Blank	0.000	03:56:14

Preventative Maintenance
Performed -- Timer Reset



EXHIBIT G-4
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST S. Hector AGENCY Prescott P.D.
DATE 3/8/16 TIME 0359
INTOXILYZER SERIAL # 80-106230 LOCATION 222 S. McIntire St.

1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

1. Clock time check.
 2. Date check.

OPERATIONAL TESTS

1. Alcohol-free subject test result 0. 000 AC.
 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
 4. Standard Calibration Check standard 0. 100 AC. Result 0. 091 AC.

Instrument is operating properly and accurately. YES NO

COMMENTS

- Annual Service - DPS
functioning properly

SIGNATURE S. Hector

INTOXILYZER 8000

Location: PRESCOTT PD

Serial Number: 80-006230

Core Version: 8105.48

03/08/2016

03:59:03

Standard Lot#: 25514100A2

Last Changed By: S. HILTON

#372

QAS: S. HILTON

#372

PRESCOTT PD

Operator: S. HILTON

#372

PRESCOTT PD

Subject: SQAP, TEST

DOB: 12/19/1973

Sex: M

Weight: 175

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	04:00:20
Diagnostic Test	Pass	04:00:55
Air Blank	0.000	04:01:23
0.100 Cal Check	0.100	04:01:43
Air Blank	0.000	04:02:13
Subject Test	0.000	04:02:33
Air Blank	0.000	04:03:00
Wait	ABT*	
Air Blank	0.000	04:03:34

*Sequence Aborted



Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: PRESCOTT PD
Serial Number: 80-006230
Core Version: 8105.48
03/08/2016

04:03:52

Standard Lot#: 25514100A2
Last Changed By: S. HILTON #372

QAS: S. HILTON #372
PRESCOTT PD

Operator: S. HILTON #372
PRESCOTT PD

Subject: SQAP, TEST
DOB: 12/19/1973
Sex: M Weight: 175

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	IPS*	04:05:13
Air Blank	0.000	04:05:41

*Improper Sample



Not a Successfully Completed Test Sequence

INTOXILYZER 8000

Location: PRESCOTT PD

Serial Number: 80-006230

Core Version: 8105.48

03/08/2016

04:06:09

Standard Lot#: 25514100A2

Last Changed By: S. HILTON #372

QAS: S. HILTON #372

PRESCOTT PD

Operator: S. HILTON #372

PRESCOTT PD

Subject: SQAP, TEST

DOB: 12/19/1973

Sex: M Weight: 175

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	RFI*	04:07:00
Air Blank	0.000	04:07:29

*RFI Detect



Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: PRESCOTT PD
Serial Number: 80-006230
Core Version: 8105.48
03/08/2016

04:07:47

Standard Lot#: 25514100A2
Last Changed By: S. HILTON #372

QAS: S. HILTON #372
PRESCOTT PD

Operator: S. HILTON #372
PRESCOTT PD

Subject: SQAP, TEST
DOB: 12/19/1973
Sex: M Weight: 175

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	04:08:50
Diagnostic Test	Pass	04:09:25
Air Blank	0.000	04:09:52
0.100 Cal Check	0.099	04:10:13
Air Blank	0.000	04:10:42
Subject Test	0.154*	04:13:50
Air Blank	0.000	04:14:24
Air Blank	0.000	04:14:52
Subject Test	XXX**	04:15:13
Air Blank	0.000	04:16:03

*Deficient Sample

**Invalid Sample

Begin new deprivation period



Not a Successfully Completed Test Sequence

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST S. Heston AGENCY Prescott P.D.
DATE 3/8/10 TIME 0916
INTOXILYZER SERIAL # 80-006230 LOCATION 222 S. MARINA ST

- (S) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.
OR
Pour a standard alcohol concentration solution .100 AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- (S) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (S) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (S) 4. Air blank completed.
- (S) 5. Calibration check completed. Test results 0. 091 AC.
- (S) 6. Air blank completed.
- (S) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE S. Heston

DPS Form Exh G-2 (Rev 05-01)

INTOXILYZER 8000

Location: PRESCOTT PD

Serial Number: 80-006230

Core Version: 8105.48

03/08/2016

04:16:43

Standard Lot#: 25514100A2

Last Changed By: S. HILTON #372

QAS: S. HILTON #372

PRESCOTT PD

Operator: S. HILTON #372

PRESCOTT PD

Test g/210L Time

Air Blank 0.000 04:17:15
0.100 Cal Check 0.091 04:17:35
Air Blank 0.000 04:18:05

Preventative Maintenance
Performed -- Timer Reset

