

INTOXILYZER 8000

Location: PRESCOTT PD

Serial Number: 80-006230

Core Version: 8105.48

12/30/2016

12:03:32

Standard Lot#: 27016100A1

Last Changed By: S. HILTON #372

QAS: S. HILTON #372

PRESCOTT PD

Operator: S. HILTON #372

PRESCOTT PD

Test	g/210L	Time
Air Blank	0.000	12:04:05
0.100 Cal Check	0.100	12:04:26
Air Blank	0.000	12:04:55

Preventative Maintenance
Performed -- Timer Reset