

EXHIBIT G-2

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R-13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST W. Lopez 30 AGENCY SEDONA P.O.

DATE 05-30-13 TIME 2330

INTOXILYZER SERIAL # 80-000924 LOCATION SEDONA P.O.

- () 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 100 AC.
 Pour a standard alcohol concentration solution N/A AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- () 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- () 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- () 4. Air blank complete.
- () 5. Calibration check complete. Test results 0.096 AC.
- () 6. Air blank complete.
- () 7. Remove printed record. Attach the record to the complete checklist.

SIGNATURE W Lopez 30

DPS Form Exb G-2 (Rev 05-01)

PASS

Intox. WAS Locked DUE TO BEING OUT OF THE 31004 TEST,
PUT INTOX. BACK INTO SERVICE.

INTOXILYZER 8000
Location: SEDONA
Serial Number: 80-000924
Core Version: 8105.48
05/30/2013 23:30:58

Standard Lot#: 09312100A1
Last Changed By: W. HUNT #461

QAS: W. LOPEZ #30
SEDONA PD

Operator: W. LOPEZ #30
SEDONA PD

Test	g/210L	Time
Air Blank	0.000	23:31:31
0.100 Cal Check	0.096	23:31:52
Air Blank	0.000	23:32:22

Preventative Maintenance
Performed -- Timer Reset