

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R-13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST Hurst AGENCY SEDONA P.D.

DATE 7-25-13 TIME 13:47 HRS

INTOXILYZER SERIAL # 80-000924 LOCATION SEDONA P.D.

- (✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution . 2 AC.
 Pour a standard alcohol concentration solution . ^{OR} N/A AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C
- (✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (✓) 4. Air blank complete.
- (✓) 5. Calibration check complete. Test results 0. 096 AC.
- (✓) 6. Air blank complete.
- (✓) 7. Remove printed record. Attach the record to the complete checklist.

SIGNATURE W. Hurst #461

DPS Form Exb G-2 (Rev 05-01)

I hereby certify that the above and forgoing is a true and correct copy of the record of periodic maintenance for the Intoxilyzer 8000 referenced therein maintained by the Sedona Police Department pursuant to the requirements of the Arizona Department of Public Safety.

Date this 25th day of July 2013

W. Hurst #461
Quality Assurance Specialist

INTOXILYZER 8000
Location: SEDONA
Serial Number: 80-000924
Core Version: 8105.48
07/25/2013 13:47:07

Standard Lot#: 09312100A1
Last Changed By: W. HUNT #461

QAS: W. HUNT #461
SEDONA PD

Operator: W. HUNT #461
SEDONA PD

Test	g/210L	Time
Air Blank	0.000	13:47:47
0.100 Cal Check	0.096	13:48:08
Air Blank	0.000	13:48:37

Preventative Maintenance
Performed -- Timer Reset

I hereby certify that the above and forgoing is a true and correct copy of the record of periodic maintenance for the intoxilyzer 8000 referenced therein maintained by the Sedona Police Department pursuant to the requirements of the Arizona Department of Public Safety.

Date this 25th day of July, 2013

W. Hunt #461
Quality Assurance Specialist

SQAP
90 DAY

EXHIBIT G-4
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST W. Hunt AGENCY Sedona PD
DATE 7-25-13 TIME 13:53 hrs
INTOXILYZER SERIAL # 80-000924 LOCATION Sedona PD

(✓) 1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS
(✓) 1. Clock time check.
(✓) 2. Date check.

OPERATIONAL TESTS
(✓) 1. Alcohol-free subject test result 0. 000 AC.
(✓) 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
(✓) 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
(✓) 4. Standard Calibration Check standard 0. 100 AC. Result 0. 096 AC.

Instrument is operating properly and accurately. YES NO

COMMENTS

SIGNATURE W. Hunt #461

DPS Form Exh G-4 (Rev 05-01)

I hereby certify that the above and foregoing is a true and correct copy of the record of periodic performance for the Intoxilyzer 8000 instrument as maintained by the Arizona Department of Public Safety to the requirements of the Arizona Department of Public Safety.

Executed 25th day of July 2013
W. Hunt #461
Quality Assurance Specialist

INTOXILYZER 8000
Location: SEDONA
Serial Number: 80-000924
Core Version: 8105.48
07/25/2013

13:53:03

Standard Lot#: 09312100A1
Last Changed By: W. HUNT #461

QAS: W. HUNT #461
SEDONA PD

Operator: W. HUNT #461
SEDONA PD

Subject: SOAP, TEST
DOB: 08/16/1964
Sex: M Weight: 250

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	13:54:48
Diagnostic Test	Pass	13:55:22
Air Blank	0.000	13:55:49
0.100 Cal Check	0.096	13:56:10
Air Blank	0.000	13:56:39
Subject Test	0.000	13:56:56
Air Blank	0.000	13:57:24
Wait	ABT*	
Air Blank	0.000	13:58:55

*Sequence Aborted

I hereby certify that the above and forgoing is a true and correct copy of the record of periodic maintenance for the Intoxilyzer 8000 referenced therein maintained by the Sedona Police Department pursuant to the requirements of the Arizona Department of Public Safety.

Date this 25th day of July, 2013

W. Hunt #461
Quality Assurance Specialist

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: SEDONA
Serial Number: 80-000924
Core Version: 8105.48
07/25/2013

13:59:41

Standard Lot#: 09312100A1
Last Changed By: W. HUNT #461

QAS: W. HUNT #461
SEDONA PD

Operator: W. HUNT #461
SEDONA PD

Subject: SQAP, TEST
DOB: 08/16/1964
Sex: M Weight: 250

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	IPS*	14:00:42
Air Blank	0.000	14:01:11

*Improper Sample

I hereby certify that the above and forgoing is a true and correct copy of the record of periodic maintenance for the Intoxilyzer 8000 referenced therein maintained by the Sedona Police Department pursuant to the requirements of the Arizona Department of Public Safety.

Date this 25th day of July, 2013

W. Hunt #461
Quality Assurance Specialist

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: SEDONA
Serial Number: 80-000924
Core Version: 8105.48
07/25/2013

14:01:45

Standard Lot#: 09312100A1
Last Changed By: W. HUNT #461

QAS: W. HUNT #461
SEDONA PD

Operator: W. HUNT #461
SEDONA PD

Subject: SOAP, TEST
DOB: 08/16/1964
Sex: M Weight: 250

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	RFI*	14:02:56
Air Blank	0.000	14:03:24

*RFI Detect

I hereby certify that the above and forgoing is a true and correct copy of the record of periodic maintenance for the intoxilyzer 8000 referenced therein maintained by the Sedona Police Department pursuant to the requirements of the Arizona Department of Public Safety.

Date this 25th day of July, 2013

W. Hunt 461
Quality Assurance Specialist

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: SEDONA
Serial Number: 80-000924
Core Version: 8105.48
07/25/2013

14:04:24

Standard Lot#: 09312100A1
Last Changed By: W. HUNT #461

QAS: W. HUNT #461
SEDONA PD

Operator: W. HUNT #461
SEDONA PD

Subject: SQAP, TEST
DOB: 08/16/1964
Sex: M Weight: 250

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	14:05:55
Diagnostic Test	Pass	14:06:29
Air Blank	0.000	14:06:57
0.100 Cal Check	0.097	14:07:17
Air Blank	0.000	14:07:47
Subject Test	0.226*	14:10:54
Air Blank	0.000	14:11:28
Air Blank	0.000	14:11:55
Subject Test	XXX**	14:12:18
Air Blank	0.000	14:12:52

*Deficient Sample

**Invalid Sample

Begin new deprivation period

I hereby certify that the above and forgoing is a true and correct copy of the record of periodic maintenance for the Intoxilyzer 8000 referenced therein maintained by the Sedona Police Department pursuant to the requirements of the Arizona Department of Public Safety.

Date this 25th day of July, 2013

W. Hunt 461
Quality Assurance Specialist

Not a Successfully Completed Test Sequence

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R-13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST Hunt AGENCY SEDONA P.D.

DATE 7-25-13 TIME _____

INTOXILYZER SERIAL # 80-000924 LOCATION SEDONA P.D.

- (✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution . 1 AC.
OR
Pour a standard alcohol concentration solution . N/A AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C
- (✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (✓) 4. Air blank complete.
- (✓) 5. Calibration check complete. Test results 0. 095 AC.
- (✓) 6. Air blank complete.
- (✓) 7. Remove printed record. Attach the record to the complete checklist.

SIGNATURE W. Hunt 461

DPS Form Exb G-2 (Rev 05-01)

I hereby certify that the above and forgoing is a true and correct copy of the record of periodic maintenance for the intoxilyzer 8000 referenced therein maintained by the Sedona Police Department pursuant to the requirements of the Arizona Department of Public Safety.

Date this 25th day of July, 2013

W. Hunt 461
Quality Assurance Specialist

INTOXILYZER 8000
 Location: SEDONA
 Serial Number: 30-000924
 Core Version: 8105.48
 07/25/2013 14:15:25

 Standard Lot#: 09312100A1
 Last Changed By: W. HUNT #461

QAS: W. HUNT #461
 SEDONA PD

Operator: W. HUNT #461
 SEDONA PD

Test	g/210L	Time
Air Blank	0.000	14:15:56
0.100 Cal Check	0.095	14:16:16
Air Blank	0.000	14:16:46

Preventative Maintenance
 Performed -- Timer Reset

I hereby certify that the above and forgoing is a true and correct copy of the record of periodic maintenance for the intoxilyzer 8000 referenced therein maintained by the Sedona Police Department pursuant to the requirements of the Arizona Department of Public Safety.

Date this 25th day of July, 2013

W. Hunt #461
 Quality Assurance Specialist