

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R-13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST Hunt 461 AGENCY SEDONA PD

DATE 02/17/15 TIME 1608

INTOXILYZER SERIAL # 80-000924 LOCATION 100 ROADRUNNER DR

(X) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution . 100 AC.

Pour a standard alcohol concentration solution . ^{OR} N/A AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C± 0.2°C

(✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"

(✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.

(✓) 4. Air blank complete.

(✓) 5. Calibration check complete. Test results 0. 097 AC.

(✓) 6. Air blank complete.

(✓) 7. Remove printed record. Attach the record to the complete checklist.

SIGNATURE W. Hunt #461

DPS Form Exb G-2 (Rev 05-01)

I hereby certify that the above and foregoing is a true and correct copy of the record of periodic maintenance for the Intoxilyzer 8000 referenced therein maintained by the Sedona Police Department pursuant to the requirements of the Arizona Department of Public Safety.

Date this 17 day of February 2015

W. Hunt 461
Quality Assurance Specialist

INTOXILYZER 8000
Location: SEDONA
Serial Number: 80-000924
Core Version: 8105.48
02/17/2015 16:08:28

Standard Lot#: 01513100A1
Last Changed By: K. HUDSPETH #474

QAS: W. HUNT #461
SEDONA PD

Operator: W. HUNT #461
SEDONA PD

Test	g/210L	Time
Air Blank	0.000	16:09:10
0.100 Cal Check	0.097	16:09:31
Air Blank	0.000	16:10:00

Preventative Maintenance
Performed -- Timer Reset

I hereby certify that the above and foregoing is a true and correct copy of the record of periodic maintenance for the Intoxilyzer 8000 referenced therein maintained by the Sedona Police Department pursuant to the requirements of the Arizona Department of Public Safety.

Date this 17 day of FEBRUARY 2015

W. Hunt 461
Quality Assurance Specialist