

EXHIBIT G-4  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R-13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY  
STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE-PROCEDURE

QA SPECIALIST W. Lopez 30 AGENCY SEDONA PD.  
DATE 06/03/15 TIME 1746  
INTOXILYZER SERIAL # 80-000924 LOCATION SEDONA PD.

(  ) 1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

- (  ) 1. Clock time check.
- (  ) 2. Date check.

OPERATIONAL TESTS

- (  ) 1. Alcohol-free subject test result 0. 000 AC.
- (  ) 2. Error recognition logic system functioning.  
Not a Successfully Completed Test Sequence printed.
- (  ) 3. Proper sample recognition system.  
Not a Successfully Completed Test Sequence printed.  
Deficient sample printed.
- (  ) 4. Standard Calibration Check standard 0. 100 AC. Result 0. 093 AC.

Instrument is operating properly and accurately. YES  NO

COMMENTS

FIRST CAL. CHECK DONE AT 1741 HOURS,  
SQAP DONE AT 1746 HOURS  
SECOND CAL. CHECK DONE AT 18  
RETURN INTOX. BACK TO SERVICE

SIGNATURE *W. Lopez* 30

DPS Form Exh G-4 (Rev 05-01)

I hereby certify that the above and forgoing is a true and correct copy of the record of periodic maintenance for the intoxilyzer 8000 referenced therein maintained by the Sedona Police Department pursuant to the requirements of the Arizona Department of Public Safety.

Dated this 03 day of JUNE, 2015.

*W. Lopez* 30  
Quality Assurance Specialist

INTOXILYZER 8000  
Location: SEDONA  
Serial Number: 80-000924  
Core Version: 8105.48  
06/03/2015 17:46:19

Standard Lot#: 14014100A1  
Last Changed By: W. LOPEZ #30

QAS: W. LOPEZ *WLP* #30  
SEDONA PD

Operator: W. LOPEZ #30  
SEDONA PD

Subject: SQAP, TEST  
DOB: 05/05/1965  
Sex: M Weight: 250

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	17:47:30
Diagnostic Test	Pass	17:48:05
Air Blank	0.000	17:48:32
0.100 Cal Check	0.098	17:48:53
Air Blank	0.000	17:49:23
Subject Test	0.000	17:49:53
Air Blank	0.000	17:50:20
Wait	ABT*	
Air Blank	0.000	17:51:07

\*Sequence Aborted

- 1 ALCONIZ FREE
- 2 START TEST BUTTON ERROR

**Not a Successfully Completed Test Sequence**

INTOXILYZER 8000  
Location: SEDONA  
Serial Number: 80-000924  
Core Version: 8105.48  
06/03/2015 17:52:25

Standard Lot#: 14014100A1  
Last Changed By: W. LOPEZ #30

QAS: W. LOPEZ *W* #30  
SEDONA PD

Operator: W. LOPEZ #30  
SEDONA PD

Subject: SQAP, TEST  
DOB: 05/05/1965  
Sex: M Weight: 250

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	IPS*	17:53:50
Air Blank	0.000	17:54:19

\*Improper Sample

*3 Blow AT lowest TIME*

**Not a Successfully Completed Test Sequence**

INTOXILYZER 8000  
Location: SEDONA  
Serial Number: 80-000924  
Core Version: 8105.48  
06/03/2015 17:58:03

Standard Lot#: 14014100A1  
Last Changed By: W. LOPEZ #30

QAS: W. LOPEZ *WJ* #30  
SEDONA PD

Operator: W. LOPEZ #30  
SEDONA PD

Subject: SQAP, TEST  
DOB: 05/05/1965  
Sex: M Weight: 250

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	RFI*	17:58:50
Air Blank	0.000	17:59:19

\*RFI Detect

4 RADIO FREQUENCY INTERFERENCE

**Not a Successfully Completed Test Sequence**

INTOXILYZER 8000

Location: SEDONA

Serial Number: 80-000924

Core Version: 8105.48

06/03/2015

18:15:50

Standard Lot#: 14014100A1

Last Changed By: W. LOPEZ #30

QAS: W. LOPEZ *WLP* #30

SEDONA PD

Operator: W. LOPEZ #30

SEDONA PD

Subject: SQAP, TEST

DOB: 05/05/1965

Sex: M

Weight: 250

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	18:17:01
Diagnostic Test	Pass	18:17:36
Air Blank	0.000	18:18:04
0.100 Cal Check	0.097	18:18:24
Air Blank	0.000	18:18:54
Subject Test	0.030*	18:22:01
Air Blank	0.000	18:22:34
Air Blank	0.000	18:23:02
Subject Test	XXX**	18:23:17
Air Blank	0.000	18:23:52

\*Deficient Sample

\*\*Invalid Sample

Begin new deprivation period

5 MOUTH ALCOHOL INCOMPLETE BLOW

6 MOUTH ALCOHOL INVALID SAMPLE

**Not a Successfully Completed Test Sequence**