

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R-13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST W. Lopez 30 AGENCY Sedona P.D.
DATE 12/07/2015 TIME 2238
INTOXILYZER SERIAL # 80-000924 LOCATION Sedona P.D.

- () 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution . 100 AC.
Pour a standard alcohol concentration solution . N/A AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34 °C± 0.2 °C
- () 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- () 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- () 4. Air blank complete.
- () 5. Calibration check complete. Test results 0. 100 AC.
- () 6. Air blank complete.
- () 7. Remove printed record. Attach the record to the complete checklist.

SIGNATURE W. Lopez 30

DPS Form Exh G-2 (Rev 05-01)

PASS

I hereby certify that the above and forgoing is a true and correct copy of the record of periodic maintenance for the intoxilyzer 8000 referenced therein maintained by the Sedona Police Department pursuant to the requirements of the Arizona Department of Public Safety.

Dated this 07 day of Dec., 2015.

W. Lopez 30
Quality Assurance Specialist

INTOXILYZER 8000
Location: SEDONA
Serial Number: 80-000924
Core Version: 8105.48
12/07/2015

22:40:19

Standard Lot#: 14014100A1
Last Changed By: W. LOPEZ #30

QAS: W. LOPEZ #30
SEDONA PD

Operator: W. LOPEZ #30 *WJ*
SEDONA PD

Test	g/210L	Time
Air Blank	0.000	22:40:52
0.100 Cal Check	0.100	22:41:13
Air Blank	0.000	22:41:42

Preventative Maintenance
Performed -- Timer Reset