

EXHIBIT G-2

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R-13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST W. Lopez 30 AGENCY Sedona P.O.  
DATE 12/14/2016 TIME 1119  
INTOXILYZER SERIAL # 80-000924 LOCATION Sedona P.O.

() 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution . 100 AC.

Pour a standard alcohol concentration solution . N/A <sup>OR</sup> AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C

() 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"

() 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.

() 4. Air blank complete.

() 5. Calibration check complete. Test results 0. 099 AC.

() 6. Air blank complete.

() 7. Remove printed record. Attach the record to the complete checklist.

SIGNATURE [Signature] 30

DPS Form Exb G-2 (Rev 05-01)

I hereby certify that the above and forgoing is a true and correct copy of the record of periodic maintenance for the intoxilyzer 8000 referenced therein maintained by the Sedona Police Department pursuant to the requirements of the Arizona Department of Public Safety.

Dated this 14 day of DEC, 2016.

[Signature] 30  
Quality Assurance Specialist

2 of 2

INTOXILYZER 8000  
Location: SEDONA  
Serial Number: 80-000924  
Core Version: 8105.48  
12/14/2016 11:19:25

-----  
Standard Lot#: 33814100A1  
Last Changed By: K. HUDSPETH #474  
-----

QAS: W. LOPEZ #30 *WLP*  
SEDONA PD  
-----

Operator: W. LOPEZ #30  
SEDONA PD  
-----

Test	g/210L	Time
Air Blank	0.000	11:20:04
0.100 Cal Check	0.099	11:20:25
Air Blank	0.000	11:20:55

Preventative Maintenance  
Performed -- Timer Reset

PASS