

EXHIBIT G-2

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R-13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000


STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST W. Lopez 30 AGENCY SEDONA P.O.

DATE 01-24-2018 TIME 0719

INTOXILYZER SERIAL # 80-000924 LOCATION SEDONA P.O.

- () 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution . 100% AC.
Pour a standard alcohol concentration solution . ^{OR} N/A AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C± 0.2°C
- () 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- () 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- () 4. Air blank complete.
- () 5. Calibration check complete. Test results 0. 099% AC.
- () 6. Air blank complete.
- () 7. Remove printed record. Attach the record to the complete checklist.

SIGNATURE  30

DPS Form Exb G-2 (Rev 05-01)

PASS

I hereby certify that the above and forgoing is a true and correct copy of the record of a periodic maintenance for the intoxilyzer 8000 referenced therein maintained by the Sedona Police Department pursuant to the requirements of the Arizona Department of Public Safety.

Dated this 24th day of Jan., 2018

Quality Assurance Specialist

W. Lopez #30

2 of 2

INTOXILYZER 8000
Location: SEDONA
Serial Number: 80-000924
Core Version: 8105.48
01/24/2018 07:19:01

Standard Lot#: 02617100A3
Last Changed By: K. HUDSPETH #474

QAS: W. LOPEZ #30 *WLL*
SEDONA PD

Operator: W. LOPEZ #30
SEDONA PD

Test	g/210L	Time
Air Blank	0.000	07:19:35
0.100 Cal Check	0.099	07:19:56
Air Blank	0.000	07:20:26

Preventative Maintenance
Performed -- Timer Reset

PKS