

BEFORE SQAP  
1 OF 2

EXHIBIT G-2

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R-13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST KEVIN HOSPEETH AGENCY SEDONA POLICE DEPARTMENT

DATE 08/01/2018 TIME 1459

INTOXILYZER SERIAL # 80-000924 LOCATION SEDONA P.D.

- (KH) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution. 100 AC.  
 Pour a standard alcohol concentration solution. N/A AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach  $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- (KH) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (KH) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (KH) 4. Air blank complete.
- (KH) 5. Calibration check complete. Test results 0. 098 AC.
- (KH) 6. Air blank complete.
- (KH) 7. Remove printed record. Attach the record to the complete checklist.

SIGNATURE [Signature] #474 PASSED

DPS Form Exb G-2 (Rev 05-01)

I hereby certify that the above and foregoing is a true and correct copy of the record of periodic maintenance for the intoxilyzer 8000 referenced therein maintained by the Sedona Police Department pursuant to the requirements of the Arizona Department of Public Safety.

Date this 1ST day of AUGUST, 2018

[Signature] #474  
Quality Assurance Specialist

BEFORE SQAP  
2 of 2

INTOXILYZER 8000  
Location: SEDONA  
Serial Number: 80-000924  
Core Version: 8105.48  
08/01/2018

14:59:14

-----  
Standard Lot#: 3491710A1  
Last Changed By: W. LOPEZ #30  
-----

QAS: K. HUDSPETH #474  
SEDONA PD  
-----

Operator: K. HUDSPETH #474  
SEDONA PD  
-----

Test	g/210L	Time
Air Blank	0.000	15:01:57
0.100 Cal Check	0.098	15:02:18
Air Blank	0.000	15:02:47

*K. Hudspeth #474* PASSED

Preventative Maintenance  
Performed -- Timer Reset

I hereby certify that the above and foregoing is a true and correct copy of the record of periodic maintenance for the Intoxilyzer 8000 referenced therein maintained by the Sedona Police Department pursuant to the requirements of the Arizona Department of Public Safety.

Date this 1<sup>ST</sup> day of AUGUST 2018

*K. Hudspeth #474*  
Quality Assurance Specialist

SQAP  
1 of 4

EXHIBIT G-4  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY  
STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000  
STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST KEVIN HUDSPETH AGENCY SEDONA POLICE DEPARTMENT  
DATE 08/01/2018 TIME 1509  
INTOXILYZER SERIAL # 80-000924 LOCATION SEDONA P.D.

(KH) 1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

- (KH) 1. Clock time check.
- (KH) 2. Date check.

OPERATIONAL TESTS

- (KH) 1. Alcohol-free subject test result 0. 000 AC.
- (KH) 2. Error recognition logic system functioning.  
Not a Successfully Completed Test Sequence printed
- (KH) 3. Proper sample recognition system.  
Not a Successfully Completed Test Sequence printed  
Deficient sample printed.
- (KH) 4. Standard Calibration Check standard 0. 100 AC. Result 0. 097 AC.

Instrument is operating properly and accurately. YES X NO     

COMMENTS

SQAP: PASSED ALL OPERATIONAL TESTS

SIGNATURE

[Signature] #474

PASSED

DPS Form Exh G-4 (Rev 05-01)

I hereby verify that the above and foregoing is a true and correct copy of the record of periodic maintenance for the Intoxilyzer 8000 referenced therein maintained by the Sedona Police Department pursuant to the requirements of the Arizona Department of Public Safety.

Date this 1<sup>ST</sup> day of AUGUST, 2018  
[Signature] #474  
Quality Assurance Specialist

SQAP  
2 of 4

INTOXILYZER 8000  
Location: SEDONA  
Serial Number: 80-000924  
Core Version: 8105.48  
08/01/2018

15:09:32

Standard Lot#: 3491710A1  
Last Changed By: W. LOPEZ #30

QAS: K. HUDSPETH #474  
SEDONA PD

Operator: K. HUDSPETH #474  
SEDONA PD

Subject: SQAP, TEST  
DOB: 04/13/1967  
Sex: M Weight: 200

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	15:11:14
Diagnostic Test	Pass	15:11:48
Air Blank	0.000	15:12:15
0.100 Cal Check	0.099	15:12:36
Air Blank	0.000	15:13:05
Subject Test	0.000	15:13:24
Air Blank	0.000	15:13:51
Wait	ABT*	
Air Blank	0.000	15:14:26

\*Sequence Aborted

ALCOHOL FREE / SEQUENCE ABORT TEST

*[Signature]* #474

**PASSED**  
I hereby certify that the above and forgoing is a true and correct copy of the record of periodic maintenance for the intoxilyzer 8000 referenced therein maintained by the Sedona Police Department pursuant to the requirements of the Arizona Department of Public Safety.

Date this 1<sup>ST</sup> day of AUGUST, 2018

*[Signature]* #474  
Quality Assurance Specialist

Not a Successfully Completed Test Sequence

SQAP  
3 of 4

INTOXILYZER 8000  
Location: SEDONA  
Serial Number: 80-000924  
Core Version: 8105.48  
08/01/2018

15:16:06

Standard Lot#: 3491710A1  
Last Changed By: W. LOPEZ #30

QAS: K. HUDSPETH #474  
SEDONA PD

Operator: K. HUDSPETH #474  
SEDONA PD

Subject: SQAP, TEST  
DOB: 04/13/1967  
Sex: M Weight: 200

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	RFI*	15:17:13
Air Blank	0.000	15:17:41

\*RFI Detect

RAD20 FREQUENCY INTERFERENCE TEST

PASSED

*K. Hudspeth* #474

I hereby certify that the above and forgoing is a true and correct copy of the record of periodic maintenance for the intoxilyzer 8000 referenced therein maintained by the Sedona Police Department pursuant to the requirements of the Arizona Department of Public Safety.

Date this 1<sup>ST</sup> day of AUGUST, 2018  
*[Signature]* #474  
Quality Assurance Specialist

Not a Successfully Completed Test Sequence

INTOXILYZER 8000  
Location: SEDONA  
Serial Number: 80-000924  
Core Version: 8105.48  
08/01/2018

SQAP  
4 of 4

15:29:28

Standard Lot#: 3491710A1  
Last Changed By: W. LOPEZ #30

OAS: K. HUDSPETH #474  
SEDONA PD

Operator: K. HUDSPETH #474  
SEDONA PD

Subject: SQAP, TEST  
DOB: 04/13/1967  
Sex: M Weight: 200

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	15:30:49
Diagnostic Test	Pass	15:31:23
Air Blank	0.000	15:31:51
0.100 Cal Check	0.098	15:32:11
Air Blank	0.000	15:32:41
Subject Test	0.053*	15:35:48
Air Blank	0.000	15:36:20
Air Blank	0.000	15:36:46
Subject Test	XXX**	15:38:04
Air Blank	0.000	15:38:37

\*Deficient Sample

\*\*Invalid Sample

PASSED

Begin new deprivation period

*[Signature]* #474

I hereby certify that the above and forgoing is a true and correct copy of the record of periodic maintenance for the intoxilyzer 8000 referenced therein maintained by the Sedona Police Department pursuant to the requirements of the Arizona Department of Public Safety.

Date this 1<sup>ST</sup> day of AUGUST, 2018

*[Signature]* #474  
Quality Assurance Specialist

Not a Successfully Completed Test Sequence

AFTER SQAP  
1 of 2

EXHIBIT G-2

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R-13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST KEVIN W HUDSPETH AGENCY SEDONA POLICE DEPARTMENT

DATE 08/01/2018 TIME 1540

INTOXILYZER SERIAL # 80-000924 LOCATION SEDONA P.D.

(KH) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution. 100 AC.

Pour a standard alcohol concentration solution. <sup>OR</sup> N/A AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach  $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$

(Lit) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"

(KH) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.

(KH) 4. Air blank complete.

(KH) 5. Calibration check complete. Test results 0. 97 AC.

(KH) 6. Air blank complete.

(Lit) 7. Remove printed record. Attach the record to the complete checklist.

SIGNATURE

[Signature] #474

PASSED

DPS Form Exb G-2 (Rev 05-01)

I hereby certify that the above and forgoing is a true and correct copy of the record of periodic maintenance for the intoxilyzer 8000 referenced therein maintained by the Sedona Police Department pursuant to the requirements of the Arizona Department of Public Safety.

Date this 1<sup>ST</sup> day of AUGUST, 2018

[Signature] #474  
Quality Assurance Specialist

AFTER SQAP  
2 of 2

INTOXILYZER 8000  
Location: SEDONA  
Serial Number: 80-000924  
Core Version: 8105.48  
08/01/2018

15:40:55

-----  
Standard Lot#: 3491710A1  
Last Changed By: W. LOPEZ #30  
-----

QAS: K. HUDSPETH #474  
SEDONA PD  
-----

Operator: K. HUDSPETH #474  
SEDONA PD  
-----

Test	g/210L	Time
Air Blank	0.000	15:41:27
0.100 Cal Check	0.097	15:41:48
Air Blank	0.000	15:42:17

PASSED

*[Signature]* #474

Preventative Maintenance  
Performed -- Timer Reset

**I hereby certify that the above and forgoing is a true and correct copy of the record of periodic maintenance for the intoxilyzer 8000 referenced therein maintained by the Sedona Police Department pursuant to the requirements of the Arizona Department of Public Safety.**

Date this 1<sup>ST</sup> day of AUGUST, 2018  
*[Signature]* #474  
Quality Assurance Specialist