

EXHIBIT G-2

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R-13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST K. McCallister AGENCY Sedona PD
DATE 9/27/18 TIME 14:51
INTOXILYZER SERIAL # 80-000924 LOCATION Sedona Booking

- () 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution . 100 AC.
OR
Pour a standard alcohol concentration solution . _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C± 0.2°C
- () 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- () 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- () 4. Air blank complete.
- () 5. Calibration check complete. Test results 0. 099 AC.
- () 6. Air blank complete.
- () 7. Remove printed record. Attach the record to the complete checklist.

SIGNATURE [Signature] # 991 passed

DPS Form Exb G-2 (Rev 05-01)

I hereby certify that the above and foregoing is a true and correct copy of the record of periodic maintenance for the intoxilyzer 8000 referenced therein maintained by the Sedona Police Department in accordance with the requirements of the Arizona Department of Public Safety.

Date this 1ST OCTOBER, 2018
[Signature] # 474
Quality Assurance Specialist

INTOXILYZER 8000
Location: SEDONA
Serial Number: 80-000924
Core Version: 8105.48
09/27/2018 14:51:44

Standard Lot#: 3491710A1
Last Changed By: W. LOPEZ #30

QAS: K. MCCULLAR #891
SEDONA PD

Operator: K. MCCULLAR #891
SEDONA PD

Test	g/210L	Time
Air Blank	0.000	14:52:20
0.100 Cal Check	0.099	14:52:41
Air Blank	0.000	14:53:10

Preventative Maintenance
Performed -- Timer Reset

K. McCullar #891 passed

I hereby certify that the above and forgoing is a true and correct copy of the record of periodic maintenance for the intoxilyzer 8000 referenced therein maintained by the Sedona Police Department pursuant to the requirements of the Arizona Department of Public Safety.

Date this 1ST day of OCTOBER, 2018
K. McCullar #474
Quality Assurance Specialist