

EXHIBIT G-2  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY  
STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000  
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST RICHARD J BYRNE III AGENCY AZ DPS

DATE 10/10/08 TIME 2341

INTOXILYZER SERIAL # 80-000929 DPS LOCATION WICKENBURG P.D.

- (RB) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.  
OR  
Pour a standard alcohol concentration solution N/A AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach  $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- (RB) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (RB) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (RB) 4. Air blank completed.
- (RB) 5. Calibration check completed. Test results 0.099 AC.
- (RB) 6. Air blank completed.
- (RB) 7. Remove printed record. Attach the record to the completed checklist.


SIGNATURE Richard J Byrne III 6386

INTOXILYZER 8000  
Location: WICKENBURG PD  
Serial Number: 80-000929  
10/10/2008 23:41:28

-----  
Standard Lot#: 735501G  
Last Changed By: R. BYRNE, III #6386  
-----

QAS: R. BYRNE, III #6386  
AZDPS  
-----

Test	g/210L	Time
Air Blank	0.000	23:41:55
Cal Check	0.099	23:42:14
Air Blank	0.000	23:42:43

 6386  
-----  
Operator's Signature