

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST RICHARD J. BYRNE III AGENCY AZ DPS
DATE 03/11/09 TIME 1155
INTOXILYZER SERIAL # 80-000929 DPS LOCATION WICKENBURG PD.

- (940) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.
OR
Pour a standard alcohol concentration solution N/A AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- (941) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (942) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (943) 4. Air blank completed.
- (944) 5. Calibration check completed. Test results 0.100 AC.
- (945) 6. Air blank completed.
- (946) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE Richard J. Byrne III #6386

INTOXILYZER 8000

Location: WICKENBURG PD

Serial Number: 80-000929

03/11/2009

11:55:16

Standard Lot#: 735501G

Last Changed By: R. BYRNE, III #6386

QAS: R. BYRNE, III #6386

AZDPS

Test	g/210L	Time
Air Blank	0.000	11:55:44
Cal Check	0.100	11:56:02
Air Blank	0.000	11:56:32

 #6386

Operator's Signature

EXHIBIT G-4
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST RICHARD J. BYRNE II AGENCY A2 DPS
DATE 03/11/09 TIME 1154
INTOXILYZER SERIAL # 80-000929 OPS LOCATION WICKENBURG P.D.

~~(94b)~~ 1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

- ~~(94b)~~ 1. Clock time check.
~~(94b)~~ 2. Date check.

OPERATIONAL TESTS

- ~~(94b)~~ 1. Alcohol-free subject test result 0.000 AC.
~~(94b)~~ 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
~~(94b)~~ 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
~~(94b)~~ 4. Standard Calibration Check standard 0.100 AC. Result 0.099 AC.

Instrument is operating properly and accurately. YES 98 NO _____

COMMENTS

LOT # 735501 EXP 12/28/09

SIGNATURE *Richard J. Byrne II* # 6386

INTOXILYZER 8000

Location: WICKENBURG PD

Serial Number : 80-000929

03/11/2009

11:56:45

Standard Lot#: 735501G

Last Changed By: R. BYRNE, III #6386

QAS: R. BYRNE, III #6386

AZDPS

Operator: R. BYRNE, III #6386

AZDPS

Subject: SQAP,TEST

DOB: 08/08/1975

Sex: M

Weight: 185

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	11:57:59
Diagnostic Test	Pass	11:58:26
Air Blank	0.000	11:58:53
0.100 Cal Check	0.100	11:59:12
Air Blank	0.000	11:59:42
Subject Test	0.000	12:00:09
Air Blank	ABT*	12:00:15
Air Blank	0.000	12:00:44

*Sequence Aborted

Not a Successfully Completed Test Sequence

INTOXILYZER 8000

Location: WICKENBURG PD

Serial Number : 80-000929

03/11/2009

12:01:13

Standard Lot#: 735501G

Last Changed By: R. BYRNE, III #6386

QAS: R. BYRNE, III #6386

AZDPS

Operator: R. BYRNE, III #6386

AZDPS

Subject: SQAP,TEST

DOB: 08/08/1975

Sex: M

Weight: 185

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	IPS*	12:02:25
Air Blank	RFI**	12:02:32

*Improper Sample

**RFI Detect

Not a Successfully Completed Test Sequence

INTOXILYZER 8000

Location: WICKENBURG PD

Serial Number : 80-000929

03/11/2009

12:03:05

Standard Lot#: 735501G

Last Changed By: R. BYRNE, III #6386

QAS: R. BYRNE, III #6386

AZDPS

Operator: R. BYRNE, III #6386

AZDPS

Subject: SQAP,TEST

DOB: 08/08/1975

Sex: M

Weight: 185

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	12:04:02
Diagnostic Test	Pass	12:04:30
Air Blank	0.000	12:04:57
0.100 Cal Check	0.099	12:05:16
Air Blank	0.000	12:05:45
Subject Test	0.116*	12:08:52
Air Blank	0.000	12:09:25
Air Blank	0.000	12:09:52
Subject Test	XXX**	12:10:24
Air Blank	0.000	12:10:57

*Deficient Sample

**Invalid Sample

Begin new deprivation period

Not a Successfully Completed Test Sequence

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST RICHARD J. BYRNE III AGENCY A2 DPS

DATE 03/11/09 TIME 1211

INTOXILYZER SERIAL # B0-000922 DPS LOCATION WICKENBURG P.D

- (PAB) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 100 AC.
Pour a standard alcohol concentration solution N/A ^{OR} AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- (PAB) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (PAB) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (PAB) 4. Air blank completed.
- (PAB) 5. Calibration check completed. Test results 0.099 AC.
- (PAB) 6. Air blank completed.
- (PAB) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE Richard J. Byrne III #6386

INTOXILYZER 8000

Location: WICKENBURG PD

Serial Number: 80-000929

03/11/2009

12:11:20

Standard Lot#: 735501G

Last Changed By: R. BYRNE, LII #6386

QAS: R. BYRNE, LII #6386

AZDPS

Test	g/210L	Time
Air Blank	0.000	12:11:47
Cal Check	0.099	12:12:06
Air Blank	0.000	12:12:36

Richard J. Byrne ^{LI} 6386

Operator's Signature