

EXHIBIT G-2  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY  
STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST RICHARD J BYRNE III AGENCY AZ DPS

DATE 08/09/09 TIME 0836

INTOXILYZER SERIAL # 80-000929 DPS LOCATION WICKENBURG P.D.

- (140) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 100 AC.  
OR  
Pour a standard alcohol concentration solution N/A AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach  $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- (140) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (140) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (140) 4. Air blank completed.
- (140) 5. Calibration check completed. Test results 0.09% AC.
- (140) 6. Air blank completed.
- (140) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE Richard J Byrne III 6386

INTOXILYZER 8000

Location: WICKENBURG PD

Serial Number: 80-000929

08/09/2009

08:36:04


-----  
Standard Lot#: 735501G

Last Changed By: R. BYRNE, III #6386  
-----

QAS: R. BYRNE, III #6386

AZDPS  
-----

Test	g/210L	Time
Air Blank	0.000	08:36:31
Cal Check	0.999	08:36:50
Air Blank	0.000	08:37:20



-----  
Operator's Signature