

EXHIBIT G-2  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY  
STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000  
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST RICHARD J. BYRNE AGENCY A2 DPS

DATE 01/11/10 TIME 2213

INTOXILYZER SERIAL # 80-000929 DPS LOCATION WICKENBURG PD

- (AB)
1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.  
Pour a standard alcohol concentration solution OR N/A AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach  $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
  2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
  3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
  4. Air blank completed.
  5. Calibration check completed. Test results 0.099 AC.
  6. Air blank completed.
  7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE Richard J. Byrne III #6386

INTOXILYZER 8000

Location: WICKENBURG PD

Serial Number: 80-000929

01/11/2010

22:13:01


-----  
Standard Lot#: 590854

Last Changed By: R. BYRNE, III #6386

-----  
QAS: R. BYRNE, III #6386

AZDPS  
-----

Test	g/210L	Time
Air Blank	0.000	22:13:29
Cal Check	0.099	22:13:47
Air Blank	0.000	22:14:17

 #6386  
-----  
Operator's Signature