

EXHIBIT G-2  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY  
STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000  
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST RICHARD J. BYRNE III AGENCY AZ DPS

DATE 03/22/10 TIME 1533

INTOXILYZER SERIAL # 80-000929 DPS LOCATION WICKENBURG P.D.

- RB* 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.  
OR  
Pour a standard alcohol concentration solution N/A AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach  $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- RB* 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- RB* 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- RB* 4. Air blank completed.
- RB* 5. Calibration check completed. Test results 0.100 AC.
- RB* 6. Air blank completed.
- RB* 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE Richard J. Byrne III #6386

HYDROLYZER 8000  
Location: WICKENBURG PD  
Serial Number: 80-008929  
Core Version: R105.43  
03/22/2010 15:37:09

-----  
Standard Lot#: 559854  
Last Changed by: R. BYRNE, IT: 46326

-----  
User: R. BYRNE, IT: 46326  
RBYNE

-----  
Operator: R. BYRNE, IT: 46326  
RBYNE

-----  
Test                      g/210L                      Time  
-----  
Air Blank                      0.000                      15:37:38  
0.100 Cal Check                0.100                      15:37:57  
Air Blank                      0.000                      15:38:28

Preventative Maintenance  
Performed -- Timer Reset