

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST RICHARD J. BYRNE III AGENCY AZ DPS
DATE 04/24/10 TIME 2339
INTOXILYZER SERIAL # 80-000 929 DPS LOCATION WICKIUPBLND P.D

- (1/10) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.
OR
Pour a standard alcohol concentration solution N/A AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- (1/10) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (1/10) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (1/10) 4. Air blank completed.
- (1/10) 5. Calibration check completed. Test results 0.100 AC.
- (1/10) 6. Air blank completed.
- (1/10) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE Richard J. Byrne III

DPS Form Exh G-2 (Rev 05-01)

RECEIVED

MAY 11 2010

YAVAPAI COUNTY ATTORNEY

80000000

150810235 0000
Location: MICROPHONE 10
Serial Number: 81000023
Core Version: B105.6P
01/24/2010

Operator: K. FZP00, 111-0079
AD108

Time	Weight	Time
Air Blank	0.000	211.100
0.100 Cal Check	0.100	211.100
Air Blank	0.000	211.100

Press and hold mouse
button to format -> Print menu