

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST RICHARD J. BYRNE III AGENCY AZ DPS

DATE 08/24/10 TIME 1908

INTOXILYZER SERIAL # 80-000929 DPS LOCATION WICKENBURG P.D.

- (RB) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.
OR
Pour a standard alcohol concentration solution N/A AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- (RB) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (RB) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (RB) 4. Air blank completed.
- (RB) 5. Calibration check completed. Test results 0.099 AC.
- (RB) 6. Air blank completed.
- (RB) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE Richard J. Byrne III #638V

INTOXILYZER 8000

Location: WICKENBURG PD

Serial Number: 80-000929

Core Version: 8105.48

08/24/2010

19:08:17

Standard Lot#: 590854

Last Changed By: R. BYRNE, III #6386

QAS: R. BYRNE, III #6386

AZDPS

Operator: R. BYRNE, III #6386

AZDPS

Test	g/210L	Time
Air Blank	0.000	19:08:48
0.100 Cal Check	0.099	19:09:09
Air Blank	0.000	19:09:28

Preventative Maintenance
Performed -- Timer Reset

EXHIBIT G-4
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST RICHARD J. BYRNE III AGENCY AZ DPS
DATE 08/24/10 TIME 1911
INTOXILYZER SERIAL # 80-000929 LOCATION WICKENBURG PD

~~(H)~~ 1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

- ~~(H)~~ 1. Clock time check.
~~(H)~~ 2. Date check.

OPERATIONAL TESTS

- ~~(H)~~ 1. Alcohol-free subject test result 0.000 AC.
~~(H)~~ 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
~~(H)~~ 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
~~(H)~~ 4. Standard Calibration Check standard 0.100 AC. Result 0.009 AC.

Instrument is operating properly and accurately. YES HJB NO

COMMENTS

LOT # 590854 Exp. 03/05/11

SIGNATURE Richard J. Byrne III #6386

INTOXILYZER 8000
Location: WICKENBURG PD
Serial Number: 80-000929
Core Version: 8105.48
08/24/2010

19:11:15

Standard Lot#: 590854
Last Changed By: R. BYRNE, III #6386

QAS: R. BYRNE, III #6386
AZDPS

Operator: R. BYRNE, III #6386
AZDPS

Subject: SQAP, TEST
DOB: 08/08/1968
Sex: M Weight: 185

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	19:12:17
Diagnostic Test	Pass	19:12:52
Air Blank	0.000	19:13:20
0.100 Cal Check	0.099	19:13:41
Air Blank	0.000	19:14:10
Subject Test	0.000	19:14:30
Air Blank	ABT*	19:14:36
Air Blank	0.000	19:15:04

*Sequence Aborted

Not a Successfully Completed Test Sequence

0100000000

INTOXILYZER 8000
Location: WICKENBURG PD
Serial Number: 80-000929
Core Version: 8105.48
08/24/2010

19:15:13

Standard Lot#: 590854
Last Changed By: R. BYRNE, III #6386

QAS: R. BYRNE, III #6386
AZDPS

Operator: R. BYRNE, III #6386
AZDPS

Subject: SQAP, TEST
DOB: 08/08/1968
Sex: M Weight: 185

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	IPS*	19:15:54
Air Blank	RFI**	19:16:02

*Improper Sample
**RFI Detect

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: WICKENBURG PD
Serial Number: 80-000929
Core Version: 8105.48
08/24/2010

19:16:24

Standard Lot#: 590854
Last Changed By: R. BYRNE, III #6386

QAS: R. BYRNE, III #6386
AZDPS

Operator: R. BYRNE, III #6386
AZDPS

Subject: SQAP, TEST
DOB: 08/08/1968
Sex: M Weight: 185

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	19:17:21
Diagnostic Test	Pass	19:17:56
Air Blank	0.000	19:18:23
0.100 Cal Check	0.101	19:18:44
Air Blank	0.000	19:19:13
Subject Test	0.291*	19:22:21
Air Blank	0.000	19:22:53
Air Blank	0.000	19:23:21
Subject Test	XXX**	19:23:40
Air Blank	0.000	19:24:14

*Deficient Sample

**Invalid Sample

Begin new deprivation period

Not a Successfully Completed Test Sequence

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST RICHARD J. BYRNE III AGENCY AZ DPS
DATE 08/21/10 TIME 1924
INTOXILYZER SERIAL # 80-000929 LOCATION WICKENBURG P.D.

1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.
OR
Pour a standard alcohol concentration solution NA AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
4. Air blank completed.
5. Calibration check completed. Test results 0.099 AC.
6. Air blank completed.
7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE Richard J. Byrne III #6380

INTOXILYZER 8000

Location: WICKENBURG PD

Serial Number: 80-000929

Core Version: 6105.48

08/24/2010

19:24:35

Standard Lot#: 590854

Last Changed By: R. BYRNE, III #6386

QAS: R. BYRNE, III #6386

AZDPS

Operator: R. BYRNE, III #6386

AZDPS

Test	g/210L	Time
Air Blank	0.000	19:25:06
0.100 Cal Check	0.099	19:25:27
Air Blank	0.000	19:25:56

Preventative Maintenance
Performed -- Timer Reset

2