

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST RICHARD J. BYRNE III AGENCY AZ DPS

DATE 09/04/11 TIME 1634

INTOXILYZER SERIAL # B0-000929 DPS LOCATION WICKENBURG PD

- (RJB) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution >100 AC.
OR
Pour a standard alcohol concentration solution N/A AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- (CAB) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (RJB) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (RJB) 4. Air blank completed.
- (RJB) 5. Calibration check completed. Test results 0.097 AC.
- (RJB) 6. Air blank completed.
- (RJB) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE Richard J. Byrne III

INPOXILYZER 8000

Location: WICKENBURG PD

Serial Number: 80-000929

Core Version: 8105.43

09/04/2011

16:34:12

Standard Lot#: 25319100A1

Last Changed By: R. BYRNE, III #6386

QAS: R. BYRNE III #6386

AZ DPS

Operator: R. BYRNE III #6386

AZ DPS

Test g/210L Time

Air Blank 0.000 16:34:45
0.100 Cal Check 0.097 16:35:06
Air Blank 0.000 16:35:35

Preventative Maintenance
Performed -- Timer Reset