

EXHIBIT G-2  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY  
STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000  
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST RICHARD J. BYRNE III AGENCY AZ DPS

DATE 10/05/11 TIME 0800

INTOXILYZER SERIAL # 80-000929 LOCATION WICKENBURG P.D.

- (RB) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.  
OR  
Pour a standard alcohol concentration solution NA AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach  $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- (RB) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (RB) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (RB) 4. Air blank completed.
- (RB) 5. Calibration check completed. Test results 0.08 AC.
- (RB) 6. Air blank completed.
- (RB) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE Richard J. Byrne III

INTOXILYZER 9000

Location: WICKENBURG PD

Serial Number: 80-000979

Core Version: 8105.08

10/05/2011

08:00:03

-----  
Standard Lot#: 25310100A1

Last Changed By: R. BYRNE, III #6386

-----  
QAS: R. BYRNE III #6386

AZ 026

-----  
Operator: R. BYRNE III #6386

AZ 026

-----  
Test                      q/210L                      Time  
-----  
Air Blank                      0.000                      08:00:34  
0.100 Cal Check                0.098                      08:00:55  
Air Blank                      0.000                      08:01:25

Preventative Maintenance  
Performed -- Timer Reset