

EXHIBIT G-2  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY  
STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000  
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST RICHARD J. BYRNE III AGENCY AZ DPS  
DATE 11/02/11 TIME 1109  
INTOXILYZER SERIAL # 80-000929 LOCATION WICKENBURG PD.

- (40) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.  
OR  
Pour a standard alcohol concentration solution NA AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach  $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- (11) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (11) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (11) 4. Air blank completed.
- (11) 5. Calibration check completed. Test results 0. 099 AC.
- (11) 6. Air blank completed.
- (11) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE Richard J. Byrne III # 6386

INTOXILYZER 8000

Location: WICKENBURG PD

Serial Number: 80-000929

Core Version: 8105.48

11/02/2011

11:09:15

-----  
Standard Lot#: 25310100A1

Last Changed By: R. BYRNE, III #6386

-----  
OAS: R. BYRNE III #6386

AZ DPS

-----  
Operator: R. BYRNE III #6386

AZ DPS

-----  
Test                      g/210L                      Time  
-----  
Air Blank                      0.000                      11:09:45  
0.100 Cal Check                0.099                      11:10:06  
Air Blank                      0.000                      11:10:35

Preventative Maintenance  
Performed -- Timer Reset

EXHIBIT G-4  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY  
STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000  
STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST RICHARD J. BYRNE III AGENCY AZ DPS  
DATE 11/02/11 TIME 1111  
INTOXILYZER SERIAL # 80-000929 DPS LOCATION WICKENBURG P.D.

~~RB~~ 1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS  
~~RB~~ 1. Clock time check.  
~~RB~~ 2. Date check.

OPERATIONAL TESTS  
~~RB~~ 1. Alcohol-free subject test result 0. 000 AC.  
~~RB~~ 2. Error recognition logic system functioning.  
Not a Successfully Completed Test Sequence printed  
~~RB~~ 3. Proper sample recognition system.  
Not a Successfully Completed Test Sequence printed  
Deficient sample printed.  
~~RB~~ 4. Standard Calibration Check standard 0. 100 AC. Result 0. 098 AC.

Instrument is operating properly and accurately. YES RB NO     

COMMENTS  
LOT #25310100A1 EXP 10/01/12  
\_\_\_\_\_  
\_\_\_\_\_

SIGNATURE Richard J. Byrne III #6386

INTOXILYZER 8000

Location: WICKENBURG PD

Serial Number: 80-000929

Core Version: 8105.48

11/02/2011

11:11:10

---

Standard Lot#: 25310100A1

Last Changed By: R. BYRNE, III #6386

---

QAS: R. BYRNE III #6386

AZ DPS

---

Operator: R. BYRNE III #6386

AZ DPS

---

Subject: SOAP, TEST

DOB: 08/08/1968

Sex: M

Weight: 185

---

15 Minute Deprivation Period? Yes

---

Test	g/210L	Time
Air Blank	0.000	11:12:09
Diagnostic Test	Pass	11:12:44
Air Blank	0.000	11:13:12
0.100 Cal Check	0.098	11:13:32
Air Blank	0.000	11:14:02
Subject Test	0.000	11:14:18
Air Blank	ABT*	11:14:27
Air Blank	0.000	11:14:55

\*Sequence Aborted

**Not a Successfully Completed Test Sequence**

INTOXILYZER 8000

Location: WICKENBURG PD.

Serial Number: 80-000929

Core Version: 8105.48

11/02/2011

11:15:07

---

Standard Lot#: 25310100A1

Last Changed By: R. BYRNE, III #6386

---

QAS: R. BYRNE III #6386

AZ DPS

---

Operator: R. BYRNE III #6386

AZ DPS

---

Subject: SQAP, TEST

DOB: 08/08/1968

Sex: M

Weight: 185

---

15 Minute Deprivation Period? Yes

---

Test	g/210L	Time
Air Blank	IPS*	11:15:46
Air Blank	RFI**	11:15:54

\*Improper Sample

\*\*RFI Detect

**Not a Successfully Completed Test Sequence**

INTOXILYZER 8000

Location: WICKENBURG PD

Serial Number: 80-000929

Core Version: 8105.48

11/02/2011

11:16:18

Standard Lot#: 25310100A1

Last Changed By: R. BYRNE, III #6386

QAS: R. BYRNE III #6386

AZ DPS

Operator: R. BYRNE III #6386

AZ DPS

Subject: SQAP, TEST

DOB: 08/08/1968

Sex: M

Weight: 185

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	11:17:23
Diagnostic Test	Pass	11:17:58
Air Blank	0.000	11:18:25
0.100 Cal Check	0.099	11:18:46
Air Blank	0.000	11:19:15
Subject Test	0.166*	11:22:23
Air Blank	0.000	11:22:55
Air Blank	0.000	11:23:22
Subject Test	XXX**	11:24:24
Air Blank	0.000	11:24:58

\*Deficient Sample

\*\*Invalid Sample

Begin new deprivation period

Not a Successfully Completed Test Sequence

EXHIBIT G-2  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY  
STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000  
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST RICHARD J. BYRNE III AGENCY AZ DPS

DATE 11/02/11 TIME 1125

INTOXILYZER SERIAL # 80-000929 DPS LOCATION WICKENBURG P.D.

- (PAB) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.  
OR  
Pour a standard alcohol concentration solution N/A AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach  $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- (PAB) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (PAB) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (PAB) 4. Air blank completed.
- (PAB) 5. Calibration check completed. Test results 0.098 AC.
- (PAB) 6. Air blank completed.
- (PAB) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE Richard J. Byrne III # 6386

TOXICOLYZER 3000  
Location: WICKENBURG PD  
Serial Number: 26-000929  
Core Version: 8105.48  
11/02/2011

11:25:19

-----  
Standard Lot#: 25310100A1  
Last Changed By: R. BYRNE, JII #6386  
-----

CAS: R. BYRNE JII #6386  
AZ DPS  
-----

Operator: R. BYRNE JII #6386  
AZ DPS  
-----

Test	g/2100	Time
Air Blank	0.000	11:25:49
0.100 Cal. Check	0.098	11:26:13
Air Blank	0.000	11:26:40

Preventative Maintenance  
Performed -- Timer Reset.

RECEIVED

NOV 11 2011

WICKENBURG POLICE DEPARTMENT