

EXHIBIT G-2  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY  
STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000  
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST RICHARD J. BYRNE # AGENCY AZ DES  
DATE 06/02/2012 TIME 1557  
INTOXILYZER SERIAL # 80-000929 DES LOCATION WICKENBURG P.D.

- HB* 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.  
OR  
Pour a standard alcohol concentration solution N/A AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach  $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- HB* 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- HB* 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- HB* 4. Air blank completed.
- HB* 5. Calibration check completed. Test results 0.098 AC.
- HB* 6. Air blank completed.
- HB* 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE Richard J. Byrne #6386

INTOXILYZER 8000

Location: WICKENBURG PD

Serial Number: 80-000929

Core Version: 8105.48

06/02/2012

15:57:36

-----  
Standard Lot#: 25310100A1

Last Changed By: R. BYRNE, III #6386

-----  
QAS: R. BYRNE III #6386

AZ DPS

-----  
Operator: R. BYRNE III #6386

AZ DPS

-----  
Test                      g/210L                      Time  
-----  
Air Blank                      0.000                      15:58:08  
0.100 Cal Check                0.098                      15:58:29  
Air Blank                      0.000                      15:58:59

Preventative Maintenance  
Performed -- Timer Reset