

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST RICHARD J. BYRNE III AGENCY AZ DPS

DATE 08/02/2012 TIME 1723

INTOXILYZER SERIAL # 80-000929 DPS LOCATION WICKENBURG P.D.

- RB* 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 100 AC.
OR
Pour a standard alcohol concentration solution N/A AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- RB* 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- RB* 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- RB* 4. Air blank completed.
- RB* 5. Calibration check completed. Test results 0.098 AC.
- RB* 6. Air blank completed.
- RB* 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE

Richard J. Byrne III

EXHIBIT G-4
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST REHARD J. BYRNE AGENCY AZ DPS
DATE 08/02/2012 TIME 1727
INTOXILYZER SERIAL # 80-000929 LOCATION WICKERBURG P.D.

~~AD~~ 1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

- ~~AD~~ 1. Clock time check.
~~AD~~ 2. Date check.

OPERATIONAL TESTS

- ~~AD~~ 1. Alcohol-free subject test result 0.000 AC.
~~AD~~ 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
~~AD~~ 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
~~AD~~ 4. Standard Calibration Check standard 0.100 AC. Result 0.098 AC.

Instrument is operating properly and accurately. YES AD NO _____

COMMENTS

LOT # 25310100A1 Exp 10/01/2012

SIGNATURE Richard J. Byrne #6386

INTOXILYZER 8000
Location: WICKENBURG PD
Serial Number: 80-000929
Core Version: 8105.48
08/02/2012

17:27:02

Standard Lot#: 25310100A1
Last Changed By: R. BYRNE, III #6386

OAS: R. BYRNE III #6386
AZ DPS

Operator: R. BYRNE III #6386
AZ DPS

Subject: SQAP, TEST
DOB: 08/08/1968
Sex: M Weight: 185

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	17:28:04
Diagnostic Test	Pass	17:28:39
Air Blank	0.000	17:29:06
0.100 Cal Check	0.098	17:29:27
Air Blank	0.000	17:29:56
Subject Test	0.000	17:30:22
Air Blank	ABT*	17:30:28
Air Blank	0.000	17:30:57

*Sequence Aborted

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: WICKENBURG PD
Serial Number: 80-000929
Core Version: 8105.48
08/02/2012 17:31:03

Standard Lot#: 25310100A1
Last Changed By: R. BYRNE, III #6386

QAS: R. BYRNE III #6386
AZ DPS

Operator: R. BYRNE III #6386
AZ DPS

Subject: SQAP, TEST
DOB: 08/08/1968
Sex: M Weight: 185

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	IPS*	17:31:51
Air Blank	RFI**	17:31:58

*Improper Sample
**RFI Detect

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: WICKENBURG PD
Serial Number: 80-000929
Core Version: 8105.48
08/02/2012 17:32:16

Standard Lot#: 25310100A1
Last Changed By: R. BYRNE, III #6386

QAS: R. BYRNE III #6386
AZ DPS

Operator: R. BYRNE III #6386
AZ DPS

Subject: SOAP, TEST
DOB: 08/08/1968
Sex: M Weight: 185

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	17:33:16
Diagnostic Test	Pass	17:33:51
Air Blank	0.000	17:34:18
0.100 Cal Check	0.098	17:34:39
Air Blank	0.000	17:35:09
Subject Test	0.130*	17:38:16
Air Blank	0.000	17:38:48
Air Blank	0.000	17:39:15
Subject Test	XXX**	17:40:19
Air Blank	0.000	17:40:53

*Deficient Sample

**Invalid Sample

Begin new deprivation period

Not a Successfully Completed Test Sequence

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST RICHARD J. BYRNE III AGENCY AZ DPS
DATE 08/21/2012 TIME 1741
INTOXILYZER SERIAL # 80-000929 DPS LOCATION WICKENBURG P.D.

- (AB) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 100 AC.
OR
Pour a standard alcohol concentration solution N/A AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- (AB) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
(AB) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
(AB) 4. Air blank completed.
(AB) 5. Calibration check completed. Test results 0.098 AC.
(AB) 6. Air blank completed.
(AB) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE Richard J. Byrne III # 6386

INTOXILYZER 8000

Location: WICKENBURG PD

Serial Number: 80-000929

Core Version: 8105.48

08/02/2012

17:41:27

Standard Lot#: 25310100A1

Last Changed By: R. BYRNE, III #6386

QAS: R. BYRNE III #6386

AZ DPS

Operator: R. BYRNE III #6386

AZ DPS

Test	g/210L	Time
Air Blank	0.000	17:42:01
0.100 Cal Check	0.098	17:42:22
Air Blank	0.000	17:42:51

Preventative Maintenance
Performed -- Timer Reset