

EXHIBIT G-2

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST RICHARD J. BYRNE III AGENCY A2 JRS

DATE 01/20/2013 TIME 1619

INTOXILYZER SERIAL # B0-000929 LOCATION WICKENBURG RD.

1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.
OR
Pour a standard alcohol concentration solution N/A AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
4. Air blank completed.
5. Calibration check completed. Test results 0.099 AC.
6. Air blank completed.
7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE Richard J. Byrne III #6386

INTOXILYZER 8000

Location: WICKENBURG PD

Serial Number: 80-000929

Core Version: 8105.48

01/20/2013

16:19:44

Standard Lot#: 15712100A1

Last Changed By: R. BYRNE III #6386

QAS: R. BYRNE III #6386

AZ DPS

Operator: R. BYRNE III #6386

AZ DPS

Test g/210L Time

Air Blank 0.000 16:20:16
0.100 Cal Check 0.099 16:20:37
Air Blank 0.000 16:21:06

Preventative Maintenance
Performed -- Timer Reset

EXHIBIT G-4
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST RICHARD J. BYRNE III AGENCY AZ DPS

DATE 01/20/2013 TIME 1623

INTOXILYZER SERIAL # 80-000929 DPS LOCATION WICKEMBURG PD.

(AB) 1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

- (AB) 1. Clock time check.
(AB) 2. Date check.

OPERATIONAL TESTS

- (AB) 1. Alcohol-free subject test result 0.000 AC.
(AB) 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
(AB) 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
(AB) 4. Standard Calibration Check standard 0.100 AC. Result 0.098 AC.

Instrument is operating properly and accurately. YES AB NO _____

COMMENTS

LOT # 15712100A1 Exp 08/01/2014

SIGNATURE Richard Byrne III #6386

INTOXILYZER 8000
Location: WICKENBURG PD
Serial Number: 80-000929
Core Version: 8105.48
01/20/2013

16:23:28

Standard Lot#: 15712100A1
Last Changed By: R. BYRNE III #6386

OAS: R. BYRNE III #6386
AZ DPS

Operator: R. BYRNE III #6386
AZ DPS

Subject: SQAP,TEST
DOB: 08/08/1968
Sex: M Weight: 185

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	16:24:24
Diagnostic Test	Pass	16:24:59
Air Blank	0.000	16:25:26
0.100 Cal Check	0.100	16:25:47
Air Blank	0.000	16:26:17
Subject Test	0.000	16:26:36
Air Blank	ABT*	16:26:42
Air Blank	0.000	16:27:11

*Sequence Aborted

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: WICKENBURG PD
Serial Number: 80-000929
Core Version: 8105.48
01/20/2013

16:27:17

Standard Lot#: 15712100A1
Last Changed By: R. BYRNE III #6386

QAS: R. BYRNE III #6386
AZ DPS

Operator: R. BYRNE III #6386
AZ DPS

Subject: SQAP,TEST
DOB: 08/08/1968
Sex: M Weight: 185

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	IPS*	16:27:52
Air Blank	RFI**	16:28:00

*Improper Sample
**RFI Detect

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: WICKENBURG PD
Serial Number: 80-000929
Core Version: 8105.48
01/20/2013 16:28:19

Standard Lot#: 15712100A1
Last Changed By: R. BYRNE III #6386

QAS: R. BYRNE III #6386
AZ DPS

Operator: R. BYRNE III #6386
AZ DPS

Subject: SQAP,TEST
DOB: 08/08/1968
Sex: M Weight: 185

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	16:29:16
Diagnostic Test	Pass	16:29:51
Air Blank	0.000	16:30:18
0.100 Cal Check	0.099	16:30:39
Air Blank	0.000	16:31:08
Subject Test	0.203*	16:34:16
Air Blank	0.000	16:34:48
Air Blank	0.000	16:35:16
Subject Test	XXX**	16:35:36
Air Blank	0.000	16:36:10

*Deficient Sample

**Invalid Sample

Begin new deprivation period

Not a Successfully Completed Test Sequence

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST RICHARD J. BYRNE III AGENCY AZ DPS
DATE 01/20/2013 TIME 1636
INTOXILYZER SERIAL # 80-000929 DPS LOCATION WICKENBURG RD.

- (1) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.
OR
Pour a standard alcohol concentration solution N/A AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- (2) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (3) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (4) 4. Air blank completed.
- (5) 5. Calibration check completed. Test results 0.098 AC.
- (6) 6. Air blank completed.
- (7) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE Richard J. Byrne III # 6386

INTOXILYZER 8000

Location: WICKENBURG PD

Serial Number: 80-000929

Core Version: 8105.48

01/20/2013

16:36:26

Standard Lot#: 15712100A1

Last Changed By: R. BYRNE III #6386

QAS: R. BYRNE III #6386

AZ DPS

Operator: R. BYRNE III #6386

AZ DPS

Test g/210L Time

Air Blank 0.000 16:36:57
0.100 Cal Check 0.098 16:37:18
Air Blank 0.000 16:37:47

Preventative Maintenance
Performed -- Timer Reset