

EXHIBIT G-2

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST RICHARD J. BYRNE III AGENCY AZ DPS

DATE 02/21/2013 TIME 1608

INTOXILYZER SERIAL # 80-000929 DPS LOCATION WICKED BULL P.D.

- (14) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.
OR
Pour a standard alcohol concentration solution NA AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- (14) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (14) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (14) 4. Air blank completed.
- (14) 5. Calibration check completed. Test results 0.098 AC.
- (14) 6. Air blank completed.
- (14) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE Richard J. Byrne III # 6386

INTOXILYZER 8000
Location: WICKENBURG PD
Serial Number: 80-000929
Core Version: 8105.48
02/21/2013 16:08:05

Standard Lot#: 15712100A1
Last Changed By: R. BYRNE III #6386

QAS: R. BYRNE III #6386
AZ DPS

Operator: R. BYRNE III #6386
AZ DPS

Test	g/210L	Time
Air Blank	0.000	16:08:39
0.100 Cal Check	0.098	16:09:00
Air Blank	0.000	16:09:29

Preventative Maintenance
Performed -- Timer Reset