

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST RICHARD J. BYRNE III AGENCY A2 DPS

DATE 04/18/2013 TIME 0820

INTOXILYZER SERIAL # 80-200929 DPS LOCATION WICKENBURG P.D.

- (48)
1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.
OR
Pour a standard alcohol concentration solution N/A AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
 4. Air blank completed.
 5. Calibration check completed. Test results 0.099 AC.
 6. Air blank completed.
 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE Richard J. Byrne III # 6386

INTOXILYZER 8000

Location: WICKENBURG PD

Serial Number: 80-000929

Core Version: 8105.48

04/18/2013

08:20:59

Standard Lot#: 15712100A1

Last Changed By: R. BYRNE III #6386

QAS: R. BYRNE III #6386

AZ DPS

Operator: R. BYRNE III #6386

AZ DPS

Test	g/210L	Time
Air Blank	0.000	08:21:43
0.100 Cal Check	0.099	08:22:04
Air Blank	0.000	08:22:33

Preventative Maintenance
Performed -- Timer Reset

EXHIBIT G-4
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST RICHARD J. BYRNE III AGENCY AZ DPS
DATE 04/18/2013 TIME 0824
INTOXILYZER SERIAL # 80-000929 DPS LOCATION WICKENBURG P.D.

1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

1. Clock time check.
 2. Date check.

OPERATIONAL TESTS

1. Alcohol-free subject test result 0. 000 AC.
 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
 4. Standard Calibration Check standard 0. 099 .100 AC. Result 0. 099 AC.

Instrument is operating properly and accurately. YES NO

COMMENTS

EXB 08/01/2014 LOT #15712100A1

SIGNATURE

Richard J. Byrne III #6306

INTOXILYZER 8000

Location: WICKENBURG PD

Serial Number: 80-000929

Core Version: 8105.48

04/18/2013

08:24:26

Standard Lot#: 15712100A1

Last Changed By: R. BYRNE III #6386

QAS: R. BYRNE III #6386

AZ DPS

Operator: R. BYRNE III #6386

AZ DPS

Subject: SOAP,TEST

DOB: 08/08/1968

Sex: M

Weight: 185

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	08:25:31
Diagnostic Test	Pass	08:26:06
Air Blank	0.000	08:26:33
0.100 Cal Check	0.100	08:26:54
Air Blank	0.000	08:27:23
Subject Test	0.000	08:27:42
Air Blank	ABT*	08:28:00
Air Blank	0.000	08:28:28

*Sequence Aborted

Not a Successfully Completed Test Sequence

INTOXILYZER 8000

Location: WICKENBURG PD

Serial Number: 80-000929

Core Version: 8105.48

04/18/2013

08:28:35

Standard Lot#: 15712100A1

Last Changed By: R. BYRNE III #6386

QAS: R. BYRNE III #6386

AZ DPS

Operator: R. BYRNE III #6386

AZ DPS

Subject: SQAP,TEST

DOB: 08/08/1968

Sex: M Weight: 185

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	IPS*	08:29:22
Air Blank	RFI**	08:29:29

*Improper Sample

**RFI Detect

Not a Successfully Completed Test Sequence

INTOXILYZER 8000

Location: WICKENBURG PD

Serial Number: 80-000929

Core Version: 8105.48

04/18/2013

08:30:02

Standard Lot#: 15712100A1

Last Changed By: R. BYRNE III #6386

QAS: R. BYRNE III #6386
AZ DPS

Operator: R. BYRNE III #6386
AZ DPS

Subject: SOAP, TEST

DOB: 08/08/1968

Sex: M Weight: 185

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	08:31:07
Diagnostic Test	Pass	08:31:42
Air Blank	0.000	08:32:10
0.100 Cal Check	0.099	08:32:31
Air Blank	0.000	08:33:00
Subject Test	0.232*	08:36:08
Air Blank	0.000	08:36:40
Air Blank	0.000	08:37:07
Subject Test	XXX**	08:37:28
Air Blank	0.000	08:38:00

*Deficient Sample

**Invalid Sample

Begin new deprivation period

Not a Successfully Completed Test Sequence

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST RICHARD J. BYRNE AGENCY A2 DPS

DATE 04/18/2013 TIME 0839

INTOXILYZER SERIAL # 80-000929 LOCATION WICKENBURG P.D.

- (AB) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.
OR
Pour a standard alcohol concentration solution N/A AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- (AB) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (AB) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (AB) 4. Air blank completed.
- (AB) 5. Calibration check completed. Test results 0.099 AC.
- (AB) 6. Air blank completed.
- (AB) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE Richard J. Byrne # 6386

INTOXILYZER 3000

Location: WICKENBURG PD

- Serial Number: 80-000929

Core Version: 8105.48

04/18/2013

08:39:55

Standard Lot#: 15712100A1

Last Changed By: R. BYRNE III #6386

CAS: R. BYRNE III #6386

AZ DPS

Operator: R. BYRNE III #6386

AZ DPS

Test g/210L Time

Air Blank 0.000 08:40:29
0.100 Cal Check 0.099 08:40:50
Air Blank 0.000 08:41:19

Preventative Maintenance
Performed -- Timer Reset