


EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST SCARIM, ROGER G. AGENCY COTTONWOOD PD
DATE 3/4/08 TIME 0753
INTOXILYZER SERIAL # 80-000933 LOCATION COTTONWOOD PD

- (M) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0.100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- (M) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
(M) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
(M) 4. Air blank completed.
(M) 5. Calibration check completed. Test results 0. 098 AC.
(M) 6. Air blank completed.
(M) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE ^{ofc}  592

DPS Form Exh G-2 (Rev 05-01)

INTOXILYZER 8000

Location: COTTONWOOD PD

Serial Number: 80-000933

03/04/2008

07:55:21

Standard Lot#: 717901E

Last Changed By: R. SCARIM #592

QAS: R. SCARIM #592

COTTONWOOD PD

Test	g/210L	Time
Air Blank	0.000	07:55:48
Cal Check	0.098	07:56:07
Air Blank	0.000	07:56:36

OR  592

Operator's Signature