


EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST SCARIM, ROGER G. AGENCY COTTONWOOD PD
DATE 6/19/09 TIME 1437
INTOXILYZER SERIAL # 80-000933 LOCATION COTTONWOOD PD

- (-) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0.100AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- (M) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
(M) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
(M) 4. Air blank completed.
(M) 5. Calibration check completed. Test results 0.099 AC.
(M) 6. Air blank completed.
(M) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE ^{OFC}  592

INTOXILYZER 8000
Location: COTTONWOOD PD
Serial Number: 80-000933
06/19/2008 14:39:38

Standard Lot#: 717901I
Last Changed By: R. SCARIM #592

QAS: R. SCARIM #592
COTTONWOOD PD

Test	g/210L	Time
Air Blank	0.000	14:40:05
Cal Check	0.099	14:40:24
Air Blank	0.000	14:40:54

ORC *[Handwritten Signature]* 592


Operator's Signature

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST SCARIM, ROGER G. AGENCY COTTONWOOD PD
DATE 6/19/08 TIME 1446
INTOXILYZER SERIAL # 80-000933 LOCATION COTTONWOOD PD

- (-1) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0.100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- (-4) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (-4) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (-4) 4. Air blank completed.
- (-4) 5. Calibration check completed. Test results 0.100 AC.
- (-4) 6. Air blank completed.
- (-4) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE ^{OR}  592

DPS Form Exh G-2 (Rev 05-01)

NOTE * BOTTLE CHANGED

08-005944


INTOXILYZER 8000
Location: COTTONWOOD PD
Serial Number: 80-000933
06/19/2008

14:46:01

Standard Lot#: 804301I
Last Changed By: R. SCARIM #592

QAS: R. SCARIM #592
COTTONWOOD PD

Test	g/210L	Time
Air Blank	0.000	14:46:28
Cal Check	0.100	14:46:47
Air Blank	0.000	14:47:16

OK  592

Operator's Signature

08-005944