

EXHIBIT G-2  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY  
STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST M. WATSON AGENCY COTTONWOOD PD  
DATE 08/28/09 TIME 1607  
INTOXILYZER SERIAL # 80-000933 LOCATION COTTONWOOD PD

- (M) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0.100 AC.  
OR  
Pour a standard alcohol concentration solution \_\_\_\_\_ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach  $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- (M) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"  
(M) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.  
(M) 4. Air blank completed.  
(M) 5. Calibration check completed. Test results 0.097 AC.  
(M) 6. Air blank completed.  
(M) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE ORC [Signature] #609

INTOXILYZER 8000  
Location: COTTONWOOD PD  
Serial Number: 80-000933  
08/28/2009

16:07:39

-----  
Standard Lot#: 585219  
Last Changed By: R. SCARIM #592  
-----

QAS: M. WATSON #609  
COTTONWOOD PD  
-----

Test	g/210L	Time
Air Blank	0.000	16:08:06
Cal Check	0.097	16:08:25
Air Blank	0.000	16:08:54


 #609  
-----  
Operator's Signature

EXHIBIT G-4  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY  
STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST M. WATSON AGENCY COTTONWOOD PD  
DATE 08/28/09 TIME 1610 HRS  
INTOXILYZER SERIAL # 80-000933 LOCATION COTTONWOOD PD

~~QAM~~ 1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

- ~~QAM~~ 1. Clock time check.  
~~QAM~~ 2. Date check.

OPERATIONAL TESTS

- ~~QAM~~ 1. Alcohol-free subject test result 0. 080 AC.  
~~QAM~~ 2. Error recognition logic system functioning.  
Not a Successfully Completed Test Sequence printed  
~~QAM~~ 3. Proper sample recognition system.  
Not a Successfully Completed Test Sequence printed  
Deficient sample printed.  
~~QAM~~ 4. Standard Calibration Check standard 0. 100 AC. Result 0. 097 AC.

Instrument is operating properly and accurately. YES  NO

COMMENTS

---

---

---

---

SIGNATURE 

INTOXILYZER 8000  
Location: COTTONWOOD PD  
Serial Number : 80-000933  
08/28/2009

16:10:14

---

Standard Lot#: 585219  
Last Changed By: R. SCARIM #592

---

QAS: M. WATSON #609  
COTTONWOOD PD

---

Operator: M. WATSON #609  
COTTONWOOD PD

---

Subject: SQAP,TEST  
DOB: 01/01/1900  
Sex: M Weight: 111

---

15 Minute Deprivation Period? Yes

---

Test	g/210L	Time
Air Blank	0.000	16:11:04
Diagnostic Test	Pass	16:11:32
Air Blank	0.000	16:11:59
0.100 Cal Check	0.097	16:12:18
Air Blank	0.000	16:12:47
Subject Test	0.000	16:13:16
Air Blank	0.000	16:13:43
Five Minute Wait	ABT*	
Air Blank	0.000	16:14:26

\*Sequence Aborted



**Not a Successfully Completed Test Sequence**

INTOXILYZER 8000  
Location: COTTONWOOD PD  
Serial Number : 80-000933  
08/28/2009

16:10:14

---

Standard Lot#: 585219  
Last Changed By: R. SCARIM #592

---

QAS: M. WATSON #609  
COTTONWOOD PD

---

Operator: M. WATSON #609  
COTTONWOOD PD

---

Subject: SQAP,TEST  
DOB: 01/01/1900  
Sex: M Weight: 111

---

15 Minute Deprivation Period? Yes

---

Test	g/210L	Time
Air Blank	0.000	16:11:04
Diagnostic Test	Pass	16:11:32
Air Blank	0.000	16:11:59
0.100 Cal Check	0.097	16:12:18
Air Blank	0.000	16:12:47
Subject Test	0.000	16:13:16
Air Blank	0.000	16:13:43
Five Minute Wait	ABT*	
Air Blank	0.000	16:14:26

\*Sequence Aborted



**Not a Successfully Completed Test Sequence**

INTOXILYZER 8000  
Location: COTTONWOOD PD  
Serial Number : 80-000933  
08/28/2009

16:10:14

---

Standard Lot#: 585219  
Last Changed By: R. SCARIM #592

---

QAS: M. WATSON #609  
COTTONWOOD PD

---

Operator: M. WATSON #609  
COTTONWOOD PD

---

Subject: SQAP, TEST  
DOB: 01/01/1900  
Sex: M Weight: 111

---

15 Minute Deprivation Period? Yes

---

Test	g/210L	Time
Air Blank	0.000	16:11:04
Diagnostic Test	Pass	16:11:32
Air Blank	0.000	16:11:59
0.100 Cal Check	0.097	16:12:18
Air Blank	0.000	16:12:47
Subject Test	0.000	16:13:16
Air Blank	0.000	16:13:43
Five Minute Wait	ABT*	
Air Blank	0.000	16:14:26

\*Sequence Aborted

  
**Not a Successfully Completed Test Sequence**

INTOXILYZER 8000  
Location: COTTONWOOD PD  
Serial Number : 80-000933  
08/28/2009

16:16:58

---

Standard Lot#: 585219  
Last Changed By: R. SCARIM #592

---

QAS: M. WATSON #609  
COTTONWOOD PD

---

Operator: M. WATSON #609  
COTTONWOOD PD

---

Subject: SQAP,TEST  
DOB: 01/01/1900  
Sex: M Weight: 111

---

15 Minute Deprivation Period? Yes

---

Test	g/210L	Time
Air Blank	IPS*	16:17:50
Air Blank	0.000	16:18:18

\*Improper Sample



**Not a Successfully Completed Test Sequence**

INTOXILYZER 8000  
Location: COTTONWOOD PD  
Serial Number : 80-000933  
08/28/2009

16:16:58

---

Standard Lot#: 585219  
Last Changed By: R. SCARIM #592

---

QAS: M. WATSON #609  
COTTONWOOD PD

---

Operator: M. WATSON #609  
COTTONWOOD PD

---

Subject: SQAP,TEST  
DOB: 01/01/1900  
Sex: M Weight: 111

---

15 Minute Deprivation Period? Yes

---

Test	g/210L	Time
Air Blank	IPS*	16:17:50
Air Blank	0.000	16:18:18

\*Improper Sample



**Not a Successfully Completed Test Sequence**

INTOXILYZER 8000  
Location: COTTONWOOD PD  
Serial Number : 80-000933  
08/28/2009

16:16:58

---

Standard Lot#: 585219  
Last Changed By: R. SCARIM #592

---

QAS: M. WATSON #609  
COTTONWOOD PD

---

Operator: M. WATSON #609  
COTTONWOOD PD

---

Subject: SQAP,TEST  
DOB: 01/01/1900  
Sex: M Weight: 111

---

15 Minute Deprivation Period? Yes

---

Test	g/210L	Time
Air Blank	IPS*	16:17:50
Air Blank	0.000	16:18:18

\*Improper Sample



**Not a Successfully Completed Test Sequence**

INTOXILYZER 8000  
Location: COTTONWOOD PD  
Serial Number : 80-000933  
08/28/2009

16:18:36

---

Standard Lot#: 585219  
Last Changed By: R. SCARIM #592

---

QAS: M. WATSON #609  
COTTONWOOD PD

---

Operator: M. WATSON #609  
COTTONWOOD PD

---

Subject: SQAP,TEST  
DOB: 01/01/1900  
Sex: M Weight: 111

---

15 Minute Deprivation Period? Yes

---

Test	g/210L	Time
Air Blank	RFI*	16:19:19
Air Blank	0.000	16:19:48

\*RFI Detect



Not a Successfully Completed Test Sequence

INTOXILYZER 8000  
Location: COTTONWOOD PD  
Serial Number : 80-000933  
08/28/2009

16:18:36

---

Standard Lot#: 585219  
Last Changed By: R. SCARIM #592

---

QAS: M. WATSON #609  
COTTONWOOD PD

---

Operator: M. WATSON #609  
COTTONWOOD PD

---

Subject: SQAP,TEST  
DOB: 01/01/1900  
Sex: M Weight: 111

---

15 Minute Deprivation Period? Yes

---

Test	g/210L	Time
Air Blank	RFI*	16:19:19
Air Blank	0.000	16:19:48

\*RFI Detect



**Not a Successfully Completed Test Sequence**

INTOXILYZER 8000  
Location: COTTONWOOD PD  
Serial Number : 80-000933  
08/28/2009

16:18:36

---

Standard Lot#: 585219  
Last Changed By: R. SCARIM #592

---

QAS: M. WATSON #609  
COTTONWOOD PD

---

Operator: M. WATSON #609  
COTTONWOOD PD

---

Subject: SQAP, TEST  
DOB: 01/01/1900  
Sex: M Weight: 111

---

15 Minute Deprivation Period? Yes

---

Test	g/210L	Time
Air Blank	RFI*	16:19:19
Air Blank	0.000	16:19:48

\*RFI Detect



**Not a Successfully Completed Test Sequence**

INTOXILYZER 8000  
Location: COTTONWOOD PD  
Serial Number : 80-000933  
08/28/2009

16:30:31

---

Standard Lot#: 585219  
Last Changed By: R. SCARIM #592

---

QAS: M. WATSON #609  
COTTONWOOD PD

---

Operator: M. WATSON #609  
COTTONWOOD PD

---

Subject: SQAP, TEST  
DOB: 01/01/1900  
Sex: M Weight: 111

---

15 Minute Deprivation Period? Yes

---

Test	g/210L	Time
Air Blank	0.000	16:31:20
Diagnostic Test	Pass	16:31:48
Air Blank	0.000	16:32:15
0.100 Cal Check	0.097	16:32:34
Air Blank	0.000	16:33:03
Subject Test	0.042*	16:36:12
Air Blank	0.000	16:36:43
Air Blank	0.000	16:37:11
Subject Test	XXX**	16:37:32
Air Blank	0.000	16:38:07

\*Deficient Sample

\*\*Invalid Sample

Begin new deprivation period

  
**Not a Successfully Completed Test Sequence**

INTOXILYZER 8000  
Location: COTTONWOOD PD  
Serial Number : 80-000933  
08/28/2009

16:30:31

---

Standard Lot#: 585219  
Last Changed By: R. SCARIM #592

---

QAS: M. WATSON #609  
COTTONWOOD PD

---

Operator: M. WATSON #609  
COTTONWOOD PD

---

Subject: SQAP,TEST  
DOB: 01/01/1900  
Sex: M Weight: 111

---

15 Minute Deprivation Period? Yes

---

Test	g/210L	Time
Air Blank	0.000	16:31:20
Diagnostic Test	Pass	16:31:48
Air Blank	0.000	16:32:15
0.100 Cal Check	0.097	16:32:34
Air Blank	0.000	16:33:03
Subject Test	0.042*	16:36:12
Air Blank	0.000	16:36:43
Air Blank	0.000	16:37:11
Subject Test	XXX**	16:37:32
Air Blank	0.000	16:38:07

\*Deficient Sample

\*\*Invalid Sample

Begin new deprivation period



**Not a Successfully Completed Test Sequence**

INTOXILYZER 8000  
Location: COTTONWOOD PD  
Serial Number : 80-000933  
08/28/2009

16:30:31

---

Standard Lot#: 585219  
Last Changed By: R. SCARIM #592

---

QAS: M. WATSON #609  
COTTONWOOD PD

---

Operator: M. WATSON #609  
COTTONWOOD PD

---

Subject: SQAP,TEST  
DOB: 01/01/1900  
Sex: M Weight: 111

---

15 Minute Deprivation Period? Yes

---

Test	g/210L	Time
Air Blank	0.000	16:31:20
Diagnostic Test	Pass	16:31:48
Air Blank	0.000	16:32:15
0.100 Cal Check	0.097	16:32:34
Air Blank	0.000	16:33:03
Subject Test	0.042*	16:36:12
Air Blank	0.000	16:36:43
Air Blank	0.000	16:37:11
Subject Test	XXX**	16:37:32
Air Blank	0.000	16:38:07

\*Deficient Sample

\*\*Invalid Sample

Begin new deprivation period



**Not a Successfully Completed Test Sequence**

EXHIBIT G-2  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY  
STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST M. WATSON AGENCY COTTONWOOD PD  
DATE 08/28/09 TIME \_\_\_\_\_  
INTOXILYZER SERIAL # 80-000933 LOCATION COTTONWOOD PD

- (Handwritten initials)* 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0.100 AC.  
OR  
Pour a standard alcohol concentration solution \_\_\_\_\_ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach  $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- (Handwritten initials)* 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"  
*(Handwritten initials)* 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.  
*(Handwritten initials)* 4. Air blank completed.  
*(Handwritten initials)* 5. Calibration check completed. Test results 0.097 AC.  
*(Handwritten initials)* 6. Air blank completed.  
*(Handwritten initials)* 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE *(Handwritten Signature)* #609

INTOXILYZER 8000

Location: COTTONWOOD PD

Serial Number: 80-000933

08/28/2009

16:39:28


-----  
Standard Lot#: 585219

Last Changed By: R. SCARIM #592

-----  
QAS: M. WATSON #609

COTTONWOOD PD  
-----

Test	g/210L	Time
Air Blank	0.000	16:39:56
Cal Check	0.097	16:40:15
Air Blank	0.000	16:40:44

OFL  #609  
-----  
Operator's Signature

