

# PROPERTY INVOICE AND RECEIPT

NOTE: FILL IN EVERY SPACE UNLESS THE INFORMATION IS NOT AVAILABLE PRINT LEGIBLY IN INK.		TYPE OF PROPERTY (Circle One Only)		DATE AND TIME PROPERTY SEIZED	
		EVIDENCE	SAFEKEEPING	FOUND	
NAME OF VICTIM OR OWNER (LAST, FIRST, MIDDLE)		ADDRESS		TELEPHONE	D.R. NUMBER
CITY OF COTTONWOOD					14-001597
NAME OF PERSON ARRESTED, OR SUSPECT (LAST, FIRST, MIDDLE)		ADDRESS		TELEPHONE	D.O.B.
NAME OF PERSON ARRESTED, OR SUSPECT (LAST, FIRST, MIDDLE)		ADDRESS		TELEPHONE	D.O.B.
NAME OF FINDER OR PERSON REPORTING FOUND PROPERTY		ADDRESS		TELEPHONE	CITATION #
CHARGES				FINDER WILL CLAIM PROPERTY	LAB TEST
				YES NO	YES NO

### NOTIFICATION FOR RETURN OF PERSONAL PROPERTY

You are responsible for claiming your personal property within 90 days from the above date. The property will be disposed of by the Cottonwood Police Department, if no claim is made prior to that date.

ADVANCE AUTHORIZATION IS REQUIRED IF YOU WISH TO DESIGNATE ANOTHER PERSON TO RECEIVE YOUR PROPERTY.

\_\_\_\_\_  
SIGNATURE

BRIEF SYNOPSIS OF OCCURRENCE: (Also use this space for additional suspects or victims)			
90 DAY SOAP			
ITEMIZE AND DESCRIBE PROPERTY			
LIST ONLY ONE ITEM PER LINE. NUMBER EACH ITEM ON THE FORM. - Where possible, place each item in plastic bag, and attach property tag to outside of bag.			
ITEM	DESCRIPTION	SERIAL #	PROPERTY #
1 <sup>st</sup>	90 DAY SOAP PAPERWORK	80-000933	92642
Placed in Property Locker #		Date	Time
		2/7/14	0840
OFFICER:	SERIAL:	PAGE	OF
off [Signature]	592	1	1
DISPOSITION OF PROPERTY:			
Property Items Were:			
1. Returned to Owner		3. Sold at Auction	
2. Destroyed		4. T.O.T. City	
5. Left at Court		6. Unable to Locate	
Received By	Date	Time	Remarks
[Signature]	2/7/14	0840	
Property Location			
DWR			

EXHIBIT G-4  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY  
STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000  
STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST SCARIM, ROGER G. AGENCY COTTONWOOD PD  
DATE 2/7/14 TIME 0805  
INTOXILYZER SERIAL # 80-000933 LOCATION COTTONWOOD PD

(M) 1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

- (M) 1. Clock time check.  
(M) 2. Date check.

OPERATIONAL TESTS

- (M) 1. Alcohol-free subject test result 0. 000 AC.  
(M) 2. Error recognition logic system functioning.  
Not a Successfully Completed Test Sequence printed  
(M) 3. Proper sample recognition system.  
Not a Successfully Completed Test Sequence printed  
Deficient sample printed.  
(M) 4. Standard Calibration Check standard 0. 100 AC. Result 0. 097 AC.

Instrument is operating properly and accurately. YES M NO     

COMMENTS

CLOCK 2 MINUTES FAST, RESET.

SIGNATURE OR M.A.R. 592

EXHIBIT G-2  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY  
STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000  
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST SCARIM, ROGER G. AGENCY COTTONWOOD PD  
DATE 2/7/14 TIME 0831  
INTOXILYZER SERIAL # 80-000933 LOCATION COTTONWOOD PD

- (M) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0.100 AC.  
OR  
Pour a standard alcohol concentration solution \_\_\_\_\_ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach  $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- (M) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (M) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (M) 4. Air blank completed.
- (M) 5. Calibration check completed. Test results 0.097 AC.
- (M) 6. Air blank completed.
- (M) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE OK M. L. 592

DPS Form Exh G-2 (Rev 05-01)

14-001597

INTOXILYZER 8000  
Location: COTTONWOOD PD  
Serial Number: 80-000933  
Core Version: 8105.48  
02/07/2014 08:12:42

Standard Lot#: 09113100A1  
Last Changed By: R. SCARIM #592

OAS: R. SCARIM #592  
COTTONWOOD PD

Operator: R. SCARIM #592  
COTTONWOOD PD

Subject: SQAP, TEST  
DOB: 11/11/1911  
Sex: M Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	IPS*	08:13:43
Air Blank	0.000	08:14:11

\*Improper Sample

Not a Successfully  
Completed  
Test Sequence

14-001597

INTOXILYZER 8000  
Location: COTTONWOOD PD  
Serial Number: 80-000933  
Core Version: 8105.48  
02/07/2014 08:06:21

Standard Lot#: 09113100A1  
Last Changed By: R. SCARIM #592

OAS: R. SCARIM #592  
COTTONWOOD PD

Operator: R. SCARIM #592  
COTTONWOOD PD

Subject: SQAP, TEST  
DOB: 11/11/1911  
Sex: M Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	08:07:50
Diagnostic Test	Pass	08:08:27
Air Blank	0.000	08:08:55
0.100 Cal Check	0.096	08:09:16
Air Blank	0.000	08:09:45
Subject Test	0.000	08:10:04
Air Blank	0.000	08:10:31
Wait	ABT*	
Air Blank	0.000	08:11:07

\*Sequence Aborted

Not a Successfully  
Completed  
Test Sequence

14-001597

INTOXILYZER 8000  
Location: COTTONWOOD PD  
Serial Number: 80-000933  
Core Version: 8105.48  
02/07/2014 08:06:21

Standard Lot#: 09113100A1  
Last Changed By: R. SCARIM #592

OAS: R. SCARIM #592  
COTTONWOOD PD

Operator: R. SCARIM #592  
COTTONWOOD PD

Subject: SQAP, TEST  
DOB: 11/11/1911  
Sex: M Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	08:07:50
Diagnostic Test	Pass	08:08:27
Air Blank	0.000	08:08:55
0.100 Cal Check	0.096	08:09:16
Air Blank	0.000	08:09:45
Subject Test	0.000	08:10:04
Air Blank	0.000	08:10:31
Wait	ABT*	
Air Blank	0.000	08:11:07

\*Sequence Aborted

Not a Successfully  
Completed  
Test Sequence

14-001597

INTOXILYZER 8000  
Location: COTTONWOOD PD  
Serial Number: 80-000933  
Core Version: 8105.48  
02/07/2014 08:17:54

Standard Lot#: 09113100A1  
Last Changed By: R. SCARIM #592

OAS: R. SCARIM #592  
COTTONWOOD PD

Operator: R. SCARIM #592  
COTTONWOOD PD

Subject: SOAP, TEST  
DOB: 11/11/1911  
Sex: M Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	RFI*	08:18:47
Air Blank	0.000	08:19:15

\*RFI Detect

Not a Successfully  
Completed  
Test Sequence

14-001597

INTOXILYZER 8000  
Location: COTTONWOOD PD  
Serial Number: 80-000933  
Core Version: 8105.48  
02/07/2014 08:12:42

Standard Lot#: 09113100A1  
Last Changed By: R. SCARIM #592

OAS: R. SCARIM #592  
COTTONWOOD PD

Operator: R. SCARIM #592  
COTTONWOOD PD

Subject: SOAP, TEST  
DOB: 11/11/1911  
Sex: M Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	IPS*	08:13:43
Air Blank	0.000	08:14:11

\*Improper Sample

Not a Successfully  
Completed  
Test Sequence

14-001597

INTOXILYZER 8000  
Location: COTTONWOOD PD  
Serial Number: 80-000933  
Core Version: 8105.48  
02/07/2014 08:12:42

Standard Lot#: 09113100A1  
Last Changed By: R. SCARIM #592

OAS: R. SCARIM #592  
COTTONWOOD PD

Operator: R. SCARIM #592  
COTTONWOOD PD

Subject: SOAP, TEST  
DOB: 11/11/1911  
Sex: M Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	IPS*	08:13:43
Air Blank	0.000	08:14:11

\*Improper Sample

Not a Successfully  
Completed  
Test Sequence

14-001597

INTOXILYZER 8000  
Location: COTTONWOOD PD  
Serial Number: 80-000933  
Core Version: 8105.48  
02/07/2014 08:22:02

Standard Lot#: 09113100A1  
Last Changed By: R. SCARIM #592

OAS: R. SCARIM #592  
COTTONWOOD PD

Operator: R. SCARIM #592  
COTTONWOOD PD

Subject: SOAP,TEST  
DOB: 11/11/1911  
Sex: M Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	08:23:06
Diagnostic Test	Pass	08:23:44
Air Blank	0.000	08:24:11
0.100 Cal Check	0.097	08:24:32
Air Blank	0.000	08:25:01
Subject Test	0.069*	08:28:09
Air Blank	0.000	08:28:41
Air Blank	0.000	08:29:08
Subject Test	XXX**	08:29:31
Air Blank	0.000	08:30:05

\*Deficient Sample

\*\*Invalid Sample

Begin new deprivation period

Not a Successfully  
Completed  
Test Sequence

14-001597

INTOXILYZER 8000  
Location: COTTONWOOD PD  
Serial Number: 80-000933  
Core Version: 8105.48  
02/07/2014 08:17:54

Standard Lot#: 09113100A1  
Last Changed By: R. SCARIM #592

OAS: R. SCARIM #592  
COTTONWOOD PD

Operator: R. SCARIM #592  
COTTONWOOD PD

Subject: SOAP,TEST  
DOB: 11/11/1911  
Sex: M Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	RFI*	08:18:47
Air Blank	0.000	08:19:15

\*RFI Detect

Not a Successfully  
Completed  
Test Sequence

14-001597

INTOXILYZER 8000  
Location: COTTONWOOD PD  
Serial Number: 80-000933  
Core Version: 8105.48  
02/07/2014 08:17:54

Standard Lot#: 09113100A1  
Last Changed By: R. SCARIM #592

OAS: R. SCARIM #592  
COTTONWOOD PD

Operator: R. SCARIM #592  
COTTONWOOD PD

Subject: SOAP,TEST  
DOB: 11/11/1911  
Sex: M Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	RFI*	08:18:47
Air Blank	0.000	08:19:15

\*RFI Detect

Not a Successfully  
Completed  
Test Sequence

14-001597

INTOXILYZER 8000  
Location: COTTONWOOD PD  
Serial Number: 80-000933  
Core Version: 8105.48  
02/07/2014 08:32:46

Standard Lot#: 09113100A1  
Last Changed By: R. SCARIM #592

QAS: R. SCARIM #592  
COTTONWOOD PD

Operator: R. SCARIM #592  
COTTONWOOD PD

Test	g/210L	Time
Air Blank	0.000	08:33:18
0.100 Cal Check	0.097	08:33:38
Air Blank	0.000	08:34:08

Preventative Maintenance  
Performed -- Timer Reset

14-001597

INTOXILYZER 8000  
Location: COTTONWOOD PD  
Serial Number: 80-000933  
Core Version: 8105.48  
02/07/2014 08:22:02

Standard Lot#: 09113100A1  
Last Changed By: R. SCARIM #592

QAS: R. SCARIM #592  
COTTONWOOD PD

Operator: R. SCARIM #592  
COTTONWOOD PD

Subject: SOAP, TEST  
DOB: 11/11/1911  
Sex: M Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	08:23:06
Diagnostic Test	Pass	08:23:44
Air Blank	0.000	08:24:11
0.100 Cal Check	0.097	08:24:32
Air Blank	0.000	08:25:01
Subject Test	0.069*	08:28:09
Air Blank	0.000	08:28:41
Air Blank	0.000	08:29:08
Subject Test	XXX**	08:29:31
Air Blank	0.000	08:30:05

\*Deficient Sample  
\*\*Invalid Sample  
Begin new deprivation period

Not a Successfully  
Completed  
Test Sequence

14-001597

INTOXILYZER 8000  
Location: COTTONWOOD PD  
Serial Number: 80-000933  
Core Version: 8105.48  
02/07/2014 08:22:02

Standard Lot#: 09113100A1  
Last Changed By: R. SCARIM #592

QAS: R. SCARIM #592  
COTTONWOOD PD

Operator: R. SCARIM #592  
COTTONWOOD PD

Subject: SOAP, TEST  
DOB: 11/11/1911  
Sex: M Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	08:23:06
Diagnostic Test	Pass	08:23:44
Air Blank	0.000	08:24:11
0.100 Cal Check	0.097	08:24:32
Air Blank	0.000	08:25:01
Subject Test	0.069*	08:28:09
Air Blank	0.000	08:28:41
Air Blank	0.000	08:29:08
Subject Test	XXX**	08:29:31
Air Blank	0.000	08:30:05

\*Deficient Sample  
\*\*Invalid Sample  
Begin new deprivation period

Not a Successfully  
Completed  
Test Sequence

14-001597