

PROPERTY INVOICE AND RECEIPT

| | | | | | |
|---|--|------------------------------------|-------------|-------------------------------|-------------|
| NOTE: FILL IN EVERY SPACE UNLESS THE INFORMATION IS NOT AVAILABLE PRINT LEGIBLY IN INK. | | TYPE OF PROPERTY (Circle One Only) | | DATE AND TIME PROPERTY SEIZED | |
| | | EVIDENCE | SAFEKEEPING | FOUND | |
| NAME OF VICTIM OR OWNER (LAST, FIRST, MIDDLE) | | ADDRESS | | TELEPHONE | D.R. NUMBER |
| CITY OF COTTONWOOD | | | | | 14-005458 |
| NAME OF PERSON ARRESTED, OR SUSPECT (LAST, FIRST, MIDDLE) | | ADDRESS | | TELEPHONE | D.O.B. |
| | | | | | |
| NAME OF PERSON ARRESTED, OR SUSPECT (LAST, FIRST, MIDDLE) | | ADDRESS | | TELEPHONE | D.O.B. |
| | | | | | |
| NAME OF FINDER OR PERSON REPORTING FOUND PROPERTY | | ADDRESS | | TELEPHONE | CITATION # |
| | | | | | |
| CHARGES | | | | FINDER WILL CLAIM PROPERTY | |
| | | | | YES | NO |
| | | | | YES | NO |

NOTIFICATION FOR RETURN OF PERSONAL PROPERTY

You are responsible for claiming your personal property within 90 days from the above date. The property will be disposed of by the Cottonwood Police Department, if no claim is made prior to that date.

ADVANCE AUTHORIZATION IS REQUIRED IF YOU WISH TO DESIGNATE ANOTHER PERSON TO RECEIVE YOUR PROPERTY.

SIGNATURE

BRIEF SYNOPSIS OF OCCURRENCE: (Also use this space for additional suspects or victims)

90 DAY SQAP

ITEMIZE AND DESCRIBE PROPERTY

LIST ONLY ONE ITEM PER LINE. NUMBER EACH ITEM ON THE FORM. - Where possible, place each item in plastic bag, and attach property tag to outside of bag.

| ITEM | DESCRIPTION | SERIAL # | PROPERTY # |
|-----------------|----------------|-----------|------------|
| 1 st | SQAP PAPERWORK | 80-000933 | 93970 |
| | | | |
| | | | |
| | | | |
| | | | |
| | | | |
| | | | |
| | | | |
| | | | |
| | | | |

| | | |
|--------------------------------------|--------------------|---------------------------|
| Placed in Property Locker # 8 | Date 5/7/14 | Time 0822 |
| OFFICER: <i>[Signature]</i> | SERIAL: 592 | PAGE 1 OF 1 |

| | | | |
|--------------------------------|---------------------|--------------------|----------------|
| DISPOSITION OF PROPERTY: | | | |
| Property Items Were: | | | |
| 1. Returned to Owner | 2. Destroyed | 3. Sold at Auction | 4. T.O.T. City |
| 5. Left at Court | 6. Unable to Locate | | |
| Received By <i>[Signature]</i> | Date 5/7/14 | Time 1400 | Remarks |
| Property Location DFW | | | |

EXHIBIT G-4
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST SCARIM, ROGER G. AGENCY COTTONWOOD PD
DATE 5/7/14 TIME 0753
INTOXILYZER SERIAL # 86-000933 LOCATION COTTONWOOD PD

(14) 1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

- (14) 1. Clock time check.
(14) 2. Date check.

OPERATIONAL TESTS

- (14) 1. Alcohol-free subject test result 0. 000 AC.
(14) 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
(14) 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
(14) 4. Standard Calibration Check standard 0. 100 AC. Result 0. 097 AC.

Instrument is operating properly and accurately. YES 14 NO _____

COMMENTS

SIGNATURE ofc [Signature] 592

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST SCAMM, ROGER G. AGENCY COTTONWOOD PD
DATE 5/7/14 TIME 0817
INTOXILYZER SERIAL # 80-000933 LOCATION COTTONWOOD PD

- (14) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0.100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- (14) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (14) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (14) 4. Air blank completed.
- (14) 5. Calibration check completed. Test results 0.097 AC.
- (14) 6. Air blank completed.
- (14) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE ofc [Signature] 592

INTOXILYZER 8000
Location: COTTONWOOD PD
Serial Number: 80-000933
Core Version: 8105.48
05/07/2014 08:00:59

Standard Lot#: 09113100A1
Last Changed By: R. SCARIM #592

QAS: R. SCARIM #592
COTTONWOOD PD

Operator: R. SCARIM #592
COTTONWOOD PD

Subject: SOAP,TEST
DOB: 11/11/1911
Sex: M Weight: 111

15 Minute Deprivation Period? Yes

| Test | g/210L | Time |
|-----------|--------|----------|
| Air Blank | IPS* | 08:02:40 |
| Air Blank | 0.000 | 08:03:08 |

*Improper Sample

Not a Successfully
Completed
Test Sequence

14-005458

INTOXILYZER 8000
Location: COTTONWOOD PD
Serial Number: 80-000933
Core Version: 8105.48
05/07/2014 07:55:30

Standard Lot#: 09113100A1
Last Changed By: R. SCARIM #592

QAS: R. SCARIM #592
COTTONWOOD PD

Operator: R. SCARIM #592
COTTONWOOD PD

Subject: SOAP,TEST
DOB: 11/11/1911
Sex: M Weight: 111

15 Minute Deprivation Period? Yes

| Test | g/210L | Time |
|-----------------|--------|----------|
| Air Blank | 0.000 | 07:56:38 |
| Diagnostic Test | Pass | 07:57:16 |
| Air Blank | 0.000 | 07:57:43 |
| 0.100 Cal Check | 0.098 | 07:58:04 |
| Air Blank | 0.000 | 07:58:33 |
| Subject Test | 0.000 | 07:58:48 |
| Air Blank | 0.000 | 07:59:15 |
| Wait | ABT* | |
| Air Blank | 0.000 | 07:59:59 |

*Sequence Aborted

Not a Successfully
Completed
Test Sequence

14-005458

INTOXILYZER 8000
Location: COTTONWOOD PD
Serial Number: 80-000933
Core Version: 8105.48
05/07/2014 07:55:30

Standard Lot#: 09113100A1
Last Changed By: R. SCARIM #592

QAS: R. SCARIM #592
COTTONWOOD PD

Operator: R. SCARIM #592
COTTONWOOD PD

Subject: SOAP,TEST
DOB: 11/11/1911
Sex: M Weight: 111

15 Minute Deprivation Period? Yes

| Test | g/210L | Time |
|-----------------|--------|----------|
| Air Blank | 0.000 | 07:56:38 |
| Diagnostic Test | Pass | 07:57:16 |
| Air Blank | 0.000 | 07:57:43 |
| 0.100 Cal Check | 0.098 | 07:58:04 |
| Air Blank | 0.000 | 07:58:33 |
| Subject Test | 0.000 | 07:58:48 |
| Air Blank | 0.000 | 07:59:15 |
| Wait | ABT* | |
| Air Blank | 0.000 | 07:59:59 |

*Sequence Aborted

Not a Successfully
Completed
Test Sequence

14-005458

INTOXILYZER 8000
Location: COTTONWOOD PD
Serial Number: 80-000933
Core Version: 8105.48
05/07/2014 07:55:30

Standard Lot#: 09113100A1
Last Changed By: R. SCARIM #592

QAS: R. SCARIM #592
COTTONWOOD PD

Operator: R. SCARIM #592
COTTONWOOD PD

Subject: SOAP,TEST
DOB: 11/11/1911
Sex: M Weight: 111

15 Minute Deprivation Period? Yes

| Test | g/210L | Time |
|-----------------|--------|----------|
| Air Blank | 0.000 | 07:56:38 |
| Diagnostic Test | Pass | 07:57:16 |
| Air Blank | 0.000 | 07:57:43 |
| 0.100 Cal Check | 0.098 | 07:58:04 |
| Air Blank | 0.000 | 07:58:33 |
| Subject Test | 0.000 | 07:58:48 |
| Air Blank | 0.000 | 07:59:15 |
| Wait | ABT* | |
| Air Blank | 0.000 | 07:59:59 |

*Sequence Aborted

Not a Successfully
Completed
Test Sequence

14-005458

INTOXILYZER 8000
Location: COTTONWOOD PD
Serial Number: 80-000933
Core Version: 8105.48
05/07/2014 08:04:11

Standard Lot#: 09113100A1
Last Changed By: R. SCARIM #592

OAS: R. SCARIM #592
COTTONWOOD PD

Operator: R. SCARIM #592
COTTONWOOD PD

Subject: SOAP,TEST
DOB: 11/11/1911
Sex: M Weight: 111

15 Minute Deprivation Period? Yes

| Test | g/210L | Time |
|-----------|--------|----------|
| Air Blank | RFI* | 08:05:53 |
| Air Blank | 0.000 | 08:06:21 |

*RFI Detect

Not a Successfully
Completed
Test Sequence

14-005458

INTOXILYZER 8000
Location: COTTONWOOD PD
Serial Number: 80-000933
Core Version: 8105.48
05/07/2014 08:04:11

Standard Lot#: 09113100A1
Last Changed By: R. SCARIM #592

OAS: R. SCARIM #592
COTTONWOOD PD

Operator: R. SCARIM #592
COTTONWOOD PD

Subject: SOAP,TEST
DOB: 11/11/1911
Sex: M Weight: 111

15 Minute Deprivation Period? Yes

| Test | g/210L | Time |
|-----------|--------|----------|
| Air Blank | RFI* | 08:05:53 |
| Air Blank | 0.000 | 08:06:21 |

*RFI Detect

Not a Successfully
Completed
Test Sequence

14-005458

INTOXILYZER 8000
Location: COTTONWOOD PD
Serial Number: 80-000933
Core Version: 8105.48
05/07/2014 08:00:59

Standard Lot#: 09113100A1
Last Changed By: R. SCARIM #592

OAS: R. SCARIM #592
COTTONWOOD PD

Operator: R. SCARIM #592
COTTONWOOD PD

Subject: SOAP,TEST
DOB: 11/11/1911
Sex: M Weight: 111

15 Minute Deprivation Period? Yes

| Test | g/210L | Time |
|-----------|--------|----------|
| Air Blank | IPS* | 08:02:40 |
| Air Blank | 0.000 | 08:03:08 |

*Improper Sample

Not a Successfully
Completed
Test Sequence

14-005458

INTOXILYZER 8000
Location: COTTONWOOD PD
Serial Number: 80-000933
Core Version: 8105.48
05/07/2014 08:00:59

Standard Lot#: 09113100A1
Last Changed By: R. SCARIM #592

OAS: R. SCARIM #592
COTTONWOOD PD

Operator: R. SCARIM #592
COTTONWOOD PD

Subject: SOAP,TEST
DOB: 11/11/1911
Sex: M Weight: 111

15 Minute Deprivation Period? Yes

| Test | g/210L | Time |
|-----------|--------|----------|
| Air Blank | IPS* | 08:02:40 |
| Air Blank | 0.000 | 08:03:08 |

*Improper Sample

Not a Successfully
Completed
Test Sequence

14-005458

INTOXILYZER 8000
Location: COTTONWOOD PD
Serial Number: 80-000933
Core Version: 8105.48
05/07/2014

08:07:24

Standard Lot#: 09113100A1
Last Changed By: R. SCARIM #592

OAS: R. SCARIM #592
COTTONWOOD PD

Operator: R. SCARIM #592
COTTONWOOD PD

Subject: SOAP,TEST
DOB: 11/11/1911
Sex: M Weight: 111

15 Minute Deprivation Period? Yes

| Test | g/210L | Time |
|-----------------|--------|----------|
| Air Blank | 0.000 | 08:08:28 |
| Diagnostic Test | Pass | 08:09:06 |
| Air Blank | 0.000 | 08:09:33 |
| 0.100 Cal Check | 0.098 | 08:09:54 |
| Air Blank | 0.000 | 08:10:23 |
| Subject Test | 0.210* | 08:13:31 |
| Air Blank | 0.000 | 08:14:03 |
| Air Blank | 0.000 | 08:14:31 |
| Subject Test | XXX** | 08:14:49 |
| Air Blank | 0.000 | 08:15:22 |

*Deficient Sample
**Invalid Sample
Begin new deprivation period

Not a Successfully
Completed
Test Sequence

14-005458

INTOXILYZER 8000
Location: COTTONWOOD PD
Serial Number: 80-000933
Core Version: 8105.48
05/07/2014

08:07:24

Standard Lot#: 09113100A1
Last Changed By: R. SCARIM #592

OAS: R. SCARIM #592
COTTONWOOD PD

Operator: R. SCARIM #592
COTTONWOOD PD

Subject: SOAP,TEST
DOB: 11/11/1911
Sex: M Weight: 111

15 Minute Deprivation Period? Yes

| Test | g/210L | Time |
|-----------------|--------|----------|
| Air Blank | 0.000 | 08:08:28 |
| Diagnostic Test | Pass | 08:09:06 |
| Air Blank | 0.000 | 08:09:33 |
| 0.100 Cal Check | 0.098 | 08:09:54 |
| Air Blank | 0.000 | 08:10:23 |
| Subject Test | 0.210* | 08:13:31 |
| Air Blank | 0.000 | 08:14:03 |
| Air Blank | 0.000 | 08:14:31 |
| Subject Test | XXX** | 08:14:49 |
| Air Blank | 0.000 | 08:15:22 |

*Deficient Sample
**Invalid Sample
Begin new deprivation period

Not a Successfully
Completed
Test Sequence

14-005458

INTOXILYZER 8000
Location: COTTONWOOD PD
Serial Number: 80-000933
Core Version: 8105.48
05/07/2014

08:07:24

Standard Lot#: 09113100A1
Last Changed By: R. SCARIM #592

OAS: R. SCARIM #592
COTTONWOOD PD

Operator: R. SCARIM #592
COTTONWOOD PD

Subject: SOAP,TEST
DOB: 11/11/1911
Sex: M Weight: 111

15 Minute Deprivation Period? Yes

| Test | g/210L | Time |
|-----------------|--------|----------|
| Air Blank | 0.000 | 08:08:28 |
| Diagnostic Test | Pass | 08:09:06 |
| Air Blank | 0.000 | 08:09:33 |
| 0.100 Cal Check | 0.098 | 08:09:54 |
| Air Blank | 0.000 | 08:10:23 |
| Subject Test | 0.210* | 08:13:31 |
| Air Blank | 0.000 | 08:14:03 |
| Air Blank | 0.000 | 08:14:31 |
| Subject Test | XXX** | 08:14:49 |
| Air Blank | 0.000 | 08:15:22 |

*Deficient Sample
**Invalid Sample
Begin new deprivation period

Not a Successfully
Completed
Test Sequence

14-005458

INTOXILYZER 8000
Location: COTTONWOOD PD
Serial Number: 80-000933
Core Version: 8105.48
05/07/2014

08:04:11

Standard Lot#: 09113100A1
Last Changed By: R. SCARIM #592

OAS: R. SCARIM #592
COTTONWOOD PD

Operator: R. SCARIM #592
COTTONWOOD PD

Subject: SOAP,TEST
DOB: 11/11/1911
Sex: M Weight: 111

15 Minute Deprivation Period? Yes

| Test | g/210L | Time |
|-----------|--------|----------|
| Air Blank | RFI* | 08:05:53 |
| Air Blank | 0.000 | 08:06:21 |

*RFI Detect

Not a Successfully
Completed
Test Sequence

14-005458

INTOXILYZER 8000
Location: COTTONWOOD PD
Serial Number: 80-000933
Core Version: 8105.48
05/07/2014 08:18:22

Standard Lot#: 09113100A1
Last Changed By: R. SCARIM #592

OAS: R. SCARIM #592
COTTONWOOD PD

Operator: R. SCARIM #592
COTTONWOOD PD

| Test | g/210L | Time |
|-----------------|--------|----------|
| Air Blank | 0.000 | 08:18:52 |
| 0.100 Cal Check | 0.097 | 08:19:13 |
| Air Blank | 0.000 | 08:19:43 |

Preventative Maintenance
Performed -- Timer Reset

14-005458