

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST SCARIN, ROGER G. AGENCY COTTONWOOD PD
DATE 10/30/15 TIME 1419
INTOXILYZER SERIAL # 80-000933 LOCATION COTTONWOOD PD

- (M) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0.100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- (M) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
(M) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
(M) 4. Air blank completed.
(M) 5. Calibration check completed. Test results 0.099 AC.
(M) 6. Air blank completed.
(M) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE ORC [Signature] 592

INTOXILYZER 8000
Instrument Initialization
13:56:12 10/30/2015

INTOXILYZER 8000
Location: COTTONWOOD PD
Serial Number: 80-000933
Core Version: 8105.48
10/30/2015 14:21:25

Standard Lot#: 32114100A1
Last Changed By: R. SCARIM #592

GAS: R. SCARIM #592
COTTONWOOD PD

Operator: R. SCARIM #592
COTTONWOOD PD

Test	g/210L	Time
Air Blank	0.000	14:21:57
0.100 Cal Check	0.099	14:22:18
Air Blank	0.000	14:22:47

Preventative Maintenance
Performed -- Timer Reset

15-012078

