

EXHIBIT G-2  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY  
STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000  
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST SCANN, ROGER G. AGENCY COTTONWOOD PD

DATE 6/28/18 TIME 1748

INTOXILYZER SERIAL # 80-000933 LOCATION COTTONWOOD PD

- (M) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0.100 AC.  
OR  
Pour a standard alcohol concentration solution \_\_\_\_\_ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach  $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- (M) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (M) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (M) 4. Air blank completed.
- (M) 5. Calibration check completed. Test results 0.098 AC.
- (M) 6. Air blank completed.
- (M) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE off. [Signature] 592

INTOXILYZER 8000  
Instrument Initialization  
17:27:39 06/28/2018

INTOXILYZER 8000  
Location: COTTONWOOD PD  
Serial Number: 80-000933  
Core Version: 8105.48  
06/28/2018 17:52:35

-----  
Standard Lot#: 02618100A1  
Last Changed By: R. SCARIM #592  
-----

QAS: R. SCARIM #592  
COTTONWOOD PD  
-----

Operator: R. SCARIM #592  
COTTONWOOD PD  
-----

Test	g/210L	Time
Air Blank	0.000	17:53:05
0.100 Cal Check	0.098	17:53:26
Air Blank	0.000	17:53:56

Preventative Maintenance  
Performed -- Timer Reset