

EXHIBIT G-4  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST SCANNIN, ROGER G. AGENCY COTTONWOOD PD  
DATE 11/29/18 TIME 1600  
INTOXILYZER SERIAL # 80-000933 LOCATION COTTONWOOD PD

1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

1. Clock time check.  
 2. Date check.

OPERATIONAL TESTS

1. Alcohol-free subject test result 0. 000 AC.  
 2. Error recognition logic system functioning.  
Not a Successfully Completed Test Sequence printed  
 3. Proper sample recognition system.  
Not a Successfully Completed Test Sequence printed  
Deficient sample printed.  
 4. Standard Calibration Check standard 0. \_\_\_\_\_ AC. Result 0. \_\_\_\_\_ AC.

Instrument is operating properly and accurately. YES \_\_\_\_\_ NO

COMMENTS

TEST NOT COMPLETED. DURING PROPER SAMPLE RECOGNITION  
INSTRUMENT INDICATED UNSTABLE SIGNAL, INTERFERENT  
DETECT AND PURGE FAIL. REMOVED FOR SERVICE

SIGNATURE ORC [Signature] JGL

INTOXILYZER 8000  
Location: COTTONWOOD PD  
Serial Number: 80-000933  
Core Version: 8105.48  
11/29/2018

16:02:00

---

Standard Lot#: 02618100A1  
Last Changed By: R. SCARIM #592

---

QAS: R. SCARIM #592  
COTTONWOOD PD

---

Operator: R. SCARIM #592  
COTTONWOOD PD

---

Subject: SQAP,TEST  
DOB: 11/11/1911  
Sex: M Weight: 111

---

15 Minute Deprivation Period? Yes

---

Test	g/210L	Time
Air Blank	0.000	16:03:16
Diagnostic Test	Pass	16:03:51
Air Blank	0.000	16:04:18
0.100 Cal Check	0.099	16:04:39
Air Blank	0.000	16:05:08
Subject Test	0.000	16:05:25
Air Blank	0.000	16:05:52
Wait	ABT*	
Air Blank	0.000	16:06:26

\*Sequence Aborted

**Not a Successfully Completed Test Sequence**

18-015215

INTOXILYZER 8000  
Location: COTTONWOOD PD  
Serial Number: 80-000933  
Core Version: 8105.48  
11/29/2018

16:02:00

---

Standard Lot#: 02618100A1  
Last Changed By: R. SCARIM #592

---

QAS: R. SCARIM #592  
COTTONWOOD PD

---

Operator: R. SCARIM #592  
COTTONWOOD PD

---

Subject: SQAP,TEST  
DOB: 11/11/1911  
Sex: M Weight: 111

---

15 Minute Deprivation Period? Yes

---

Test	g/210L	Time
Air Blank	0.000	16:03:16
Diagnostic Test	Pass	16:03:51
Air Blank	0.000	16:04:18
0.100 Cal Check	0.099	16:04:39
Air Blank	0.000	16:05:08
Subject Test	0.000	16:05:25
Air Blank	0.000	16:05:52
Wait	ABT*	
Air Blank	0.000	16:06:26

\*Sequence Aborted

Not a Successfully Completed Test Sequence

18-015215

INTOXILYZER 8000  
Location: COTTONWOOD PD  
Serial Number: 80-000933  
Core Version: 8105.48  
11/29/2018

16:02:00

---

Standard Lot#: 02618100A1  
Last Changed By: R. SCARIM #592

---

QAS: R. SCARIM #592  
COTTONWOOD PD

---

Operator: R. SCARIM #592  
COTTONWOOD PD

---

Subject: SQAP,TEST  
DOB: 11/11/1911  
Sex: M Weight: 111

---

15 Minute Deprivation Period? Yes

---

Test	g/210L	Time
Air Blank	0.000	16:03:16
Diagnostic Test	Pass	16:03:51
Air Blank	0.000	16:04:18
0.100 Cal Check	0.099	16:04:39
Air Blank	0.000	16:05:08
Subject Test	0.000	16:05:25
Air Blank	0.000	16:05:52
Wait	ABT*	
Air Blank	0.000	16:06:26

\*Sequence Aborted

**Not a Successfully Completed Test Sequence**

18-015213

INTOXILYZER 8000  
Location: COTTONWOOD PD  
Serial Number: 80-000933  
Core Version: 8105.48  
11/29/2018

16:07:06

---

Standard Lot#: 02618100A1  
Last Changed By: R. SCARIM #592

---

QAS: R. SCARIM #592  
COTTONWOOD PD

---

Operator: R. SCARIM #592  
COTTONWOOD PD

---

Subject: SQAP,TEST  
DOB: 11/11/1911  
Sex: M Weight: 111

---

15 Minute Deprivation Period? Yes

---

Test	g/210L	Time
Air Blank	IPS*	16:07:48
Air Blank	0.000	16:08:17

\*Improper Sample

Not a Successfully Completed Test Sequence

18-015215

INTOXILYZER 8000  
Location: COTTONWOOD PD  
Serial Number: 80-000933  
Core Version: 8105.48  
11/29/2018

16:07:06

---

Standard Lot#: 02618100A1  
Last Changed By: R. SCARIM #592

---

QAS: R. SCARIM #592  
COTTONWOOD PD

---

Operator: R. SCARIM #592  
COTTONWOOD PD

---

Subject: SQAP, TEST  
DOB: 11/11/1911  
Sex: M Weight: 111

---

15 Minute Deprivation Period? Yes

---

Test	g/210L	Time
Air Blank	IPS*	16:07:48
Air Blank	0.000	16:08:17

\*Improper Sample

Not a Successfully Completed Test Sequence

18-015215

INTOXILYZER 8000  
Location: COTTONWOOD PD  
Serial Number: 80-000933  
Core Version: 8105.48  
11/29/2018

16:07:06

---

Standard Lot#: 02618100A1  
Last Changed By: R. SCARIM #592

---

QAS: R. SCARIM #592  
COTTONWOOD PD

---

Operator: R. SCARIM #592  
COTTONWOOD PD

---

Subject: SQAP, TEST  
DOB: 11/11/1911  
Sex: M Weight: 111

---

15 Minute Deprivation Period? Yes

---

Test	g/210L	Time
Air Blank	IPS*	16:07:48
Air Blank	0.000	16:08:17

\*Improper Sample

Not a Successfully Completed Test Sequence

18-015215

INTOXILYZER 8000  
Location: COTTONWOOD PD  
Serial Number: 80-000933  
Core Version: 8105.48  
11/29/2018

16:08:54

---

Standard Lot#: 02618100A1  
Last Changed By: R. SCARIM #592

---

QAS: R. SCARIM #592  
COTTONWOOD PD

---

Operator: R. SCARIM #592  
COTTONWOOD PD

---

Subject: SQAP, TEST  
DOB: 11/11/1911  
Sex: M Weight: 111

---

15 Minute Deprivation Period? Yes

---

Test	g/210L	Time
Air Blank	RFI*	16:09:43
Air Blank	0.000	16:10:11

\*RFI Detect

Not a Successfully Completed Test Sequence

18-015215

INTOXILYZER 8000

Location: COTTONWOOD PD

Serial Number: 80-000933

Core Version: 8105.48

11/29/2018

16:08:54

---

Standard Lot#: 02618100A1

Last Changed By: R. SCARIM

#592

---

QAS: R. SCARIM

#592

COTTONWOOD PD

---

Operator: R. SCARIM

#592

COTTONWOOD PD

---

Subject: SQAP, TEST

DOB: 11/11/1911

Sex: M

Weight: 111

---

15 Minute Deprivation Period? Yes

---

Test	g/210L	Time
Air Blank	RFI*	16:09:43
Air Blank	0.000	16:10:11

\*RFI Detect

Not a Successfully Completed Test Sequence

18-015215

INTOXILYZER 8000  
Location: COTTONWOOD PD  
Serial Number: 80-000933  
Core Version: 8105.48  
11/29/2018 16:08:54

---

Standard Lot#: 02618100A1  
Last Changed By: R. SCARIM #592

---

QAS: R. SCARIM #592  
COTTONWOOD PD

---

Operator: R. SCARIM #592  
COTTONWOOD PD

---

Subject: SQAP,TEST  
DOB: 11/11/1911  
Sex: M Weight: 111

---

15 Minute Deprivation Period? Yes

---

Test	g/210L	Time
Air Blank	RFI*	16:09:43
Air Blank	0.000	16:10:11

\*RFI Detect

Not a Successfully Completed Test Sequence

18-015215

INTOXILYZER 8000  
Location: COTTONWOOD PD  
Serial Number: 80-000933  
Core Version: 8105.48  
11/29/2018

16:10:39

---

Standard Lot#: 02618100A1  
Last Changed By: R. SCARIM #592

---

QAS: R. SCARIM #592  
COTTONWOOD PD

---

Operator: R. SCARIM #592  
COTTONWOOD PD

---

Subject: SQAP,TEST  
DOB: 11/11/1911  
Sex: M Weight: 111

---

15 Minute Deprivation Period? Yes

---

Test	g/210L	Time
Air Blank	0.000	16:11:32
Diagnostic Test	Pass	16:12:07
Air Blank	0.000	16:12:34
0.100 Cal Check	INT*	16:12:56
Air Blank	PUR**	16:13:43

\*Interferent Detect

\*\*Purge Fail

Not a Successfully Completed Test Sequence

18-015215

INTOXILYZER 8000  
Location: COTTONWOOD PD  
Serial Number: 80-000933  
Core Version: 8105.48  
11/29/2018

16:10:39

---

Standard Lot#: 02618100A1  
Last Changed By: R. SCARIM #592

---

QAS: R. SCARIM #592  
COTTONWOOD PD

---

Operator: R. SCARIM #592  
COTTONWOOD PD

---

Subject: SQAP, TEST  
DOB: 11/11/1911  
Sex: M Weight: 111

---

15 Minute Deprivation Period? Yes

---

Test	g/210L	Time
Air Blank	0.000	16:11:32
Diagnostic Test	Pass	16:12:07
Air Blank	0.000	16:12:34
0.100 Cal Check	INT*	16:12:56
Air Blank	PUR**	16:13:43

\*Interferent Detect

\*\*Purge Fail

Not a Successfully Completed Test Sequence

18-01521J

INTOXILYZER 8000  
Location: COTTONWOOD PD  
Serial Number: 80-000933  
Core Version: 8105.48  
11/29/2018

16:10:39

---

Standard Lot#: 02618100A1  
Last Changed By: R. SCARIM #592

---

QAS: R. SCARIM #592  
COTTONWOOD PD

---

Operator: R. SCARIM #592  
COTTONWOOD PD

---

Subject: SQAP, TEST  
DOB: 11/11/1911  
Sex: M Weight: 111

---

15 Minute Deprivation Period? Yes

---

Test	g/210L	Time
Air Blank	0.000	16:11:32
Diagnostic Test	Pass	16:12:07
Air Blank	0.000	16:12:34
0.100 Cal Check	INT*	16:12:56
Air Blank	PUR**	16:13:43

\*Interferent Detect

\*\*Purge Fail

Not a Successfully Completed Test Sequence

18-015215

INTOXILYZER 8000  
Location: COTTONWOOD PD  
Serial Number: 80-000933  
Core Version: 8105.48  
11/29/2018

16:21:59

---

Standard Lot#: 02618100A1  
Last Changed By: R. SCARIM #592

---

QAS: R. SCARIM #592  
COTTONWOOD PD

---

Operator: R. SCARIM #592  
COTTONWOOD PD

---

Subject: SQAP,TEST  
DOB: 11/11/1911  
Sex: M Weight: 111

---

15 Minute Deprivation Period? Yes

---

Test	g/210L	Time
Air Blank	0.000	16:23:01
Diagnostic Test	Pass	16:23:36
Air Blank	0.000	16:24:03
0.100 Cal Check	0.099	16:24:24
Air Blank	0.000	16:24:54
Subject Test	0.127*	16:28:01
Air Blank	0.000	16:28:34
Air Blank	0.000	16:29:01
Reference	UNS**	16:29:09
Air Blank	0.000	16:29:37

\*Deficient Sample  
\*\*Unstable Signal

Not a Successfully Completed Test Sequence

18-015215

INTOXILYZER 8000  
Location: COTTONWOOD PD  
Serial Number: 80-000933  
Core Version: 8105.48  
11/29/2018

16:21:59

---

Standard Lot#: 02618100A1  
Last Changed By: R. SCARIM #592

---

QAS: R. SCARIM #592  
COTTONWOOD PD

---

Operator: R. SCARIM #592  
COTTONWOOD PD

---

Subject: SQAP,TEST  
DOB: 11/11/1911  
Sex: M Weight: 111

---

15 Minute Deprivation Period? Yes

---

Test	g/210L	Time
Air Blank	0.000	16:23:01
Diagnostic Test	Pass	16:23:36
Air Blank	0.000	16:24:03
0.100 Cal Check	0.099	16:24:24
Air Blank	0.000	16:24:54
Subject Test	0.127*	16:28:01
Air Blank	0.000	16:28:34
Air Blank	0.000	16:29:01
Reference	UNS**	16:29:09
Air Blank	0.000	16:29:37

\*Deficient Sample

\*\*Unstable Signal

**Not a Successfully Completed Test Sequence**

14-015215

INTOXILYZER 8000  
Location: COTTONWOOD PD  
Serial Number: 80-000933  
Core Version: 8105.48  
11/29/2018

16:21:59

---

Standard Lot#: 02618100A1  
Last Changed By: R. SCARIM #592

---

QAS: R. SCARIM #592  
COTTONWOOD PD

---

Operator: R. SCARIM #592  
COTTONWOOD PD

---

Subject: SQAP, TEST  
DOB: 11/11/1911  
Sex: M Weight: 111

---

15 Minute Deprivation Period? Yes

---

Test	g/210L	Time
Air Blank	0.000	16:23:01
Diagnostic Test	Pass	16:23:36
Air Blank	0.000	16:24:03
0.100 Cal Check	0.099	16:24:24
Air Blank	0.000	16:24:54
Subject Test	0.127*	16:28:01
Air Blank	0.000	16:28:34
Air Blank	0.000	16:29:01
Reference	UNS**	16:29:09
Air Blank	0.000	16:29:37

\*Deficient Sample

\*\*Unstable Signal

Not a Successfully Completed Test Sequence

18-015215